

CALIBRATION CERTIFICATE

K B DIAGNOSTICS CENTER

Vanthavasi

604 408

Dear Sir/Madam,

As desired by you please find enclosed herewith the calibration report

Instrument Model ; EM200

Serial number ; B200450

Range of the instrument : optical density Linear from 0-3.5 A

Calibration done Date : 18TH APR 2024

Calibration due Date : 17th APR 2025

Pressure	Ideal	Actual
	1 bar	0.98bar
Lamp voltage	12 V	12.31 V
Temperature	37 ° C	37 ° C
Photometer calibration check		ok
Cuvette blank check		ok
Wavelength	Gain	
340	860	
405	601	
505	524	
546	505	
578	468	
600	503	
660	480	

CALIBRATION CERTIFICATE

700	438
-----	-----

The instrument does photometer check before each batch of tests is run. It will give an error message if

The photometer test fails and it will not run any samples.

Thanking you and assuring of our prompt attention at all times.

Your faithfully,

For TRANSASIA BIO-MEDICALS LTD.