

# Ashuma Diagnostic Solutions (OPC) Pvt. Ltd.

## CALIBRATION REPORT

Calibration Certificate No:	51/2024
Calibration Date:	08.08.2024
4Calibration Due Date:	07.08.2025

Laboratory Name: Deep Clinical Laboratory

Contact Name, Mr. Shreekant Parandekar

Address: Karve Nagar, Pune 411052

### Description of Product:

Name: Temperature indicator (Fridge-1)  
Sr. No: NA  
Make: NA  
Model: NA  
Type: Digital

Range: 2 to 8°C  
Resolution: NA  
Working Range: As per Manual  
Sr. No: NA  
Location: Laboratory

### Details of Standard used for calibration:

Name: Digital Temperature mater  
Make: Mextech  
Model: M 288 CTH  
Type: Digital  
Condition: Ok

Range: -50 to 70°C  
Certificate No: ACIS/2023/02/27/001  
Certified by: Accutech Calibration,  
Calibration Validity: 26/05/2025

### Calibration Results (Temperature)

Fridge Reading displayed in °C	Standard Reading in °C	Error in °C
6.4	7.0	-0.6
-0.4	-0.6	-0.2

Remark: The value measure above is the mean of 5 Readings

End of the Certificate.

