

HORIBA

**HORIBA India Private Limited**

246, Okhla Industrial Estate Phase-II,  
New Delhi 110020, India  
Tel: +91 (11) 4646 5000 / 4669 5001  
Fax: +91 (11) 4646 5020  
http://www.horiba.com  
CIN: U73100DL2006PTC153232

HIN/MED/2023-2024 /10045625  
13<sup>th</sup> SEP 2023


## **CALIBRATION CERTIFICATE**

This is to certify that the Hematology Analyzer Horiba YUMIZEN H550 bearing serial number 207YAXH03841 installed at **KRSNAA DIAGNOSTICS LTD.- ROHRU** as calibrated on 13<sup>TH</sup> SEP 2023.

Calibrator : ABX Minocal  
Lot No. : CX436  
Expiry Date : 05<sup>th</sup> OCT 2023

The reports of Blank Cycle, Repeatability and Calibration Values were all found in acceptable range.

Next calibration cycle is due on 12<sup>TH</sup> SEP 2024



Shrish Dixit

(Head- Products & Marketing)  
For **Horiba India Pvt. Ltd.**

246, Okhla Industrial Estate Phase-II,  
New Delhi 110020, India  
Tel: +91 (11) 4646 5000 / 46695001  
Fax: +91 (11) 4646 5020  
http://www.horiba.com  
CIN: U73100DL2006PTC1M232