



# The Accurate Calibration & Medical Solution

## CALIBRATION CERTIFICATE

Page 1 of 1

CALIBRATION CERTIFICATE OF MICROSCOPE			
Certificate No :	TAC/12-23/KDL/MS-1001	Calibration Date :	18.12.2023
Job Identification No :	TAC/C/18/12/2023	Next Calibration Date :	17.12.2024
Ref. No :	SRF Dated - 16. 12. 2023	Issue Date :	20.12.2023
Customer Name	M/s Krsnaa Diagnostic Ltd.		
Address	Civil Hospital Rohru Shimla -171207		

DESCRIPTION OF THE EQUIPMENT TO BE CALIBRATED			
Instrument Name	Micro Scope	Visual Inspection	Ok
Make	Magnus	I.D. No.	KDHPROCHHM25
Model	GLOMEX	Location / Dept.	Haematology
Serial No.	22FO194		

CALIBRATION STANDARD USED						
Sl.No.	Master Equipment	Manufacturer	Model No.	Serial No.	Date of Calibration	Valid Upto
1	Electrical Safety Analyzer	DATREND Systems Inc.	vPed - Mini	VPM18100028	14.10.2023	14.10.2024

ENVIRONMENTAL CONDITIONS	Temperature	Humidity
	22.0 °C ± 2.0 °C	50 ± 10%RH

### RESULT

VISUAL TEST		Tests	Remarks
Sr.No.			OK
1	Power Cords, Cable Checking		OK
2	Mains Socket Checking		CF
3	Equipment Type (B,BF,CF)		I
4	Equipment Class ( I, II, III )		

FUNCTIONAL TEST		Parameters	Measured Value	Remarks
Sr.No.				OK
1	Voltage Between Live and Neutral (VLN)		230.4 V (220-240 V)	OK
2	Voltage Between Live and Earth (VLE)		229.2 V (220-240V)	OK
3	Voltage Between Neutral and Earth (VNE)		0.3 V (0-5V)	OK
4	Load Current (Ia)		0.4 A	OK
5	Enclosure Current		0.7 μA (0-10μA)	OK
6	Leaking Current (IL)		112 μA (0-300μA)	OK
7	Time for Testing (t)		15 (min)	OK


Electrical Safety Test Passed as per IEC Standard 60601.1

Calibration Passed YES  NO

Remarks:

1. Certificates shall not be reproduced except the written approval of The Accurate Calibration & Medical Solution.
2. The Calibration result reported in this certificate are only valid at the time of Calibration and the stated condition of measurement.

Calibrated by



Approved by

Plot No. 72/G1, Ganesh Vihar Sirsi Road, Jaipur (Raj.) -302012

Mob.: 9672294758

E-mail : theaccuratecalibration@gmail.com