




A A CALIBRATION PVT. LTD.

CALIBRATION CERTIFICATE



7.8F-01

CC - 2646

| | | | | | | | |
|----------------------|-------------------------------------|---------------------|------|----------------------------------|--------------|----------|--------|
| Certificate No | AACPL / 09174F | Field | Mech | Group | Volume | Page No. | 1 of 1 |
| SRF No & Date | 231218.1 | DT - 18 / 12 / 2023 | | ULR No. | CC2646230000 | 09174F | |
| Calibration Method : | Comparison Method | | | Accuracy | | | |
| Company Name | M/s Krishna Diagnostic Ltd | | | Calibration Date | 18.12.2023 | | |
| Address | Civil Hospital Rohru Shimla -171207 | | | Recommended Date for Next Callb. | 20.12.2024 | | |
| | | | | Certificate Issue Dt. | 20.12.2023 | | |
| | | | | UUC Condition | Good | | |

Calibration Instrument Detail

| SI | Instrument Name | Range | Least Count | Make / SI.No. | Party ID / Location |
|----|-----------------|--------------|-------------|---------------|---------------------------|
| 1 | Micro Pipette | 10 to 100 µl | 1 µl | Dragon Lab | KDH/PROCHSE20 Serology |

Standard Equipments Used (Traceable to National Standard)

| SI | Instrument Name | Make/SI No | Calibrated By | Cal Certificate No | Cal Valid Upto |
|----|------------------|-----------------|----------------|--------------------|----------------|
| 1 | Weighing Balance | Radwag / 669683 | Bluebox Techno | BBT/452/JAN/23 | 05/03/2024 |

Reference Standard ISO - 8655 - 6 - 2022

| Reference Standard | Calibration Purpose | Humidity | Temperature | Calibration Performed At |
|-----------------------|---------------------|----------|-------------|--------------------------|
| ISO - 8655 - 6 - 2022 | Volume Measurement | 54 %RH | 23.6 °C | Laboratory |

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 01C, Section No. - CP - 03

| S. No | Nominal Value on UUC (in µl) | Calibration Result | | | Uncertainty at approx 95% confidence level and coverage factor k = 2 is ± (in µl) |
|-------|------------------------------|------------------------|-------------------|--|---|
| | | Average Volume (in µl) | Deviation (in µl) | Level and coverage factor k = 2 is ± (in µl) | |
| 1 | 10 | 9.97868 | 0.02132 | 0.25 | |
| 2 | 50 | 49.96543 | 0.03457 | 0.25 | |
| 3 | 100 | 99.92127 | 0.07873 | 0.25 | |

- Note**
- The calibration results reported in this calibration certificate are valid at the time of & under stated condition
 - This certificate cannot be reproduced except in full without our prior permission in writing
 - This certificate refers only to the particular item(s) submitted for calibration
 - UUC - Unit under calibration

----- End of Certificate -----

Calibrated by *A. K. Chaurasia*



Approved by *A. K. Chaurasia*
(CEO)

LAB :- Plot No. C-18, Sector-13, First Floor, "Vasundhara, Haat" Vasundhara, Ghaziabad - 201012
Ph. : 0120-4135529, Mob. : 9212172564, 9212172597
E-mail : aacplinfo2009@gmail.com, aacplinfo@rediffmail.com