



A A CALIBRATION PVT. LTD.

CALIBRATION CERTIFICATE



CC - 2646

Certificate No	AACPL / 09175F	Field	Mech.	Group	Volume	Page No.	1 of 1
SRF No & Date	231218.1	DT - 18 / 12 / 2023		ULR No.	CC2646230000 09175F		
Calibration Method :	Comparison Method			Accuracy			
Company Name	M/s Krsnaa Diagnostic Ltd			Calibration Date	18.12.2023		
Address	Civil Hospital Rohru Shimla -171207			Recommended Date for Next Calib.	17.12.2024		
				Certificate Issue Dt.	20.12.2023		
				UUC Condition	Good		

Calibration Instrument Detail

SI	Instrument Name	Range	Least Count	Make / SI.No.	Party ID / Location
1	Micro Pipette	100 to 1000 µl	5 µl	Dragon Lab YE231BD0015570	KDHPROCHSE21 Serology

Standard Equipments Used (Traceable to National Standard)

SI	Instrument Name	Make/SI No	Calibrated By	Cal Certificate No	Cal. Valid Upto
1	Weighing Balance	Radwag / 668683	Bluebox Techno	BBT/452/JAN/23	05/03/2024

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
ISO - 8655 - 6 - 2022	Volume Measurement	54 %RH	23.2 °C	Laboratory

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 01C, Section No : CP - 03

Calibration Result


S. No	Nominal Value on UUC (in µl)	Average Volume at 27 °C (in µl)	Deviation (in µl)	Uncertainty at approx 95% confidence Level and coverage factor $k = 2$ is \pm (in µl)
1	100	100.03146	-0.03146	0.25
2	500	500.05913	-0.05913	0.25
3	1000	1000.09287	-0.09287	0.25

Note

- 1 The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- 2 This certificate cannot be reproduced except in full without our prior permission in writing
- 3 This certificate refers only to the particular item(s) submitted for calibration
- 4 UUC - Unit under calibration

----- End of Certificate -----

Asish
Calibrated by



Approved by
Anand Chaurasia
(CEO)