



PROFICIENCY TESTING REPORT
ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME
 NABL accredited program as per ISO/IEC 17043:2010 standard
 Organized By Department of Hematology, AIIMS, New Delhi-110029



Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens

EQAP CODE No. : 2640

Distribution No.: 163-F

Month/Year: March/2024

Instrument ID: HORIBA

Model Name.: YUMIZEN H 500

Serial No.: 001YOXH03179

Name & Contact No. of PT Co-ordinator: Dr. Manoranjan Mahapatra (Prof. & Head), Hematology, AIIMS, Delhi,
 Tel: 9013085730 , E-Mail : info@ishtmaiimseqap.com

Date of issue & status of the report: 10-05-2024[Final].

CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 ³ /µl	1	5.03	6.94	11.97	12.8	0.003	-1.19	1.91	0.1	0.008	-1.10
RBC x10 ⁶ /µl	1	4.24	4.19	8.43	8.41	0.007	0.10	0.05	0.04	0.002	0.27
Hb g/dl	1	13.3	13.3	26.6	26.8	0.020	-0.34	0	0.1	0.008	-0.67
HCT%	1	38.3	38.2	76.5	84.5	0.183	-1.46	0.1	0.4	0.024	-0.81
MCV-fl	1	91.1	90.4	181.5	200.9	0.360	-1.47	0.7	0.2	0.019	1.69
MCH-Pg	1	31.6	31.5	63.1	63.9	0.060	-0.49	0.1	0.3	0.018	-0.67
MCHC-g/dl	1	34.8	34.7	69.5	63.2	0.126	1.35	0.1	0.3	0.013	-0.67
Plt. x10 ³ /µl	1	211	200	411	367	1.134	1.41	11	4	0.246	1.57
Retic %	2	6.5	6.2	12.7	19	0.335	-0.59	0.3	0.6	0.044	-0.58

P.S . Assesment

YOUR REPORT			CONSENSUS REPORT		
DLC%	3	Nrbcs=0 , Poly=56 L=13, E=2, Mono/Promono=2 , B1=8 P.M.=4, Mye=10, Meta=05, Other=	Poly: 64 - 75, Lympho: 6 - 12, Myelo: 4 - 9, Meta: 2 - 6, Eosino: 2- 5, Mono: 1-5, Promyelo/Blast/Baso: 0-5		
RBC Morphology	3	Normocytic Normochromic	Predominantly: Normocytic/Normochromic; Moderate: Anisocytosis, Mild: Poikilocytosis, Macrocytes, Tear drop cells		
Diagnosis	3	CML	Chronic Myeloid Leukemia (Chronic Phase)		

COMBINED DATA VALUES OF TOTAL PARTICIPANTS

Test parameters	S.No.	Total participants covered in the current dist. 163--F	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
WBC x10 ³ /μl	1	345	341	89.15	88.27	4.4	6.74	6.45	4.99
RBC x10 ⁶ /μl	1	345	345	85.51	88.7	7.83	4.93	6.66	6.37
Hb g/dl	1	345	345	85.8	88.41	7.54	4.93	6.66	6.66
HCT%	1	345	341	94.72	91.5	3.23	4.99	2.05	3.51
MCV-fl	1	345	341	98.53	86.51	1.47	5.28	0	8.21
MCH-Pg	1	345	341	87.68	91.79	6.74	3.23	5.58	4.98
MCHC-g/dl	1	345	341	96.19	91.79	3.23	2.64	0.58	5.57
Plt. x10 ³ /μl	1	345	340	90.29	93.24	4.71	3.82	5	2.94
ReticCount%	2	345	288	94.79	82.64	3.82	11.46	1.39	5.90
PS Assessment	3	345	274	Satisfactory :73.64%, Borderline Sat. :24.83, Unsatisfactory :1.53%					

Comments:

1). Among Lab (EQA) : **Acceptable Results.**

2). Within Lab (IQA) : **Precision acceptable.**

Note-1: EQA (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

IQA (Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

Note-2: Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

Note-3: Z score 0 to ±2: Acceptable, Z score ±2 to ±3 :Warning Signal, Z score > ±3 : Unacceptable [As per ISO/IEC 13528:2015 standard]

Note-4: Z score value between "0 to ±2" are texted in green colour. Z score value between "±2 to ±3" are texted in orange colour. Z score value > ±3 are texted in red colour.

Note-5: Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value (0.3*SDPA). To pass the stability test, average difference in measurement values of first and last day sample (x-y) should be smaller than the check value (0.3*SDPA).

Note-6: ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

Note-7: Participants are free to use methods/analyzer of their own choice.

Note-8: Proficiency testing (PT) samples are sent quarterly to each participant.

Note-9: All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website www.ishtmaiimseqap.com.

Note 10: Reports are kept confidential.

Report authorized by,



Dr. Manoranjan Mahapatra (Prof. & Head)
PT Co-ordinator: ISHTM-AIIMS-EQAP
Department of Hematology, AIIMS, New Delhi

No action observed such that NOCAPA Required.



-----End Of Report-----