



Calibration Certificate

Following voltages have been checked & adjusted on **SYS 200 PRO (Sl. No. SYS202200907)** at **KRSNAA DIAGNOSTIC LIMITED BILASPUR, HP**

Calibration Report:

System voltages are in the specified range –Checked by Digital Multimeter are as below:

S.No.	Module	Observed Value	Acceptable Range
1	Power supply board o/p	+24.1 VDC	+24.0 +/- 0.5 VDC
		+05.1 VDC	+05.0 +/- 0.2 VDC
2	Lamp voltage	+12.2 VDC	12.0 +/- 0.5 VDC

System Temperature values are in the specified range –Checked by Digital Thermometer are as below:



S.No.	Module	Observed Value	Acceptable Range
1	Reaction Disk Temp	37.0 C	37.0 C +/- 0.2 C

Photometer Adjustment Values are within range as recommended:

S.No.	Module	Observed Value	Acceptable Range
1	ADC Count for all Wavelengths	19,700 to 20,200	19,500 to 20,500

- All other mechanisms have been checked through field service program & Software, found OK.
- Performed calibration for End Point, Kinetic & Fixed Time Kinetic. Run control & sample. Results were satisfactory.
- Next Calibration will be performed on 19/10/2023

For DiaSys Diagnostics India Pvt. Ltd.

Asmat Sami
RSM (North India)

Date:19/10/2022

Registered Office
Plot No. A-821, T.T.C. Industrial Area,
MIDC, Mahape, Navi Mumbai - 400710
Maharashtra, India

DiaSys Diagnostics India Private Limited
The Specialists for diagnostic system solutions
Point-of care testing and automated Laboratory systems

CIN : U74999MH2013PTC248371
Email : info@diasys.in
Tel. No. : +91-022-71068331
Toll Free : 1800 120 1447
www.diasys-diagnostics.com