



A A CALIBRATION PVT. LTD.

CALIBRATION CERTIFICATE

7.8F-01



CC - 2646

Certificate No	AACPL / 03463F	Field	Mech.	Group	Speed	Page No.	1 of 1
SRF No & Date	240507.4 DT - 07 / 05 / 2024	ULR No.	CC2646240000 03463F				
Calibration Method :	Comparison Method		Accuracy				
Company Name	M/s Krsnaa Diagnostics Ltd.			Calibration Date	07.05.2024		
Address	8QQ8+783, Regional Hospital Rd, Cinema Colony, Changer Sector, Kosrian Sector, Bilaspur, Himachal Pradesh			Recommended Date for Next Callb.	06.05.2025		
				Certificate Issue Dt.	13.05.2024		
				UUC Condition	Good		

Calibration Instrument Detail

SI	Instrument Name	Range	Least Count	Make / SI.No.	Party ID / Location
1	Card Centrifuge M/c	880 RPM	NA	Matrix BHCC-267/22-23	KDHPBLRHMM05 Hematology

Standard Equipments Used (Traceable to National Standard)

SI	Instrument Name	Make/SI No	Calibrated By	Cal Certificate No	Cal. Valid Upto
1	Dig. Tachometer	Line Seiki / -----	AACPL	AACPL / 06363F	31/08/2024

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
IS : 12508 - 2018	RPM Measurement	55 %RH	24.6 °C	Site

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 01B, Section No : CP - 08

Calibration Result

SI	Calib. Gauge Value	Master Gauge Value	Deviation	Uncertainty at approx 95% confidence Level and coverage factor $k = 2$ is \pm (in RPM)
	UUC (in RPM)	(in RPM)	(in RPM)	
1	880	879.2	0.8	3.90

Note

- The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- This certificate cannot be reproduced except in full without our prior permission in writing
- This certificate refers only to the particular item(s) submitted for calibration
- UUC - Unit under calibration

----- End of Certificate -----

As per
Calibrated by



Approved by
Anand
Anand Chaurasia
(CEO)