



Metropolis Healthcare Ltd., EQAS

KRSNAA DIAGNOSTICS,LTD,GURDASPUR (1039)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 240104 and Sample No 01

Report Date : 10/07/2024

Analyte	Instrument	Result Value	Standard Unit	Z-Score
X Uric Acid	Diasys 200 Pro	2.43	mg/dl	-4.86
Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : Diasys 200 Pro


Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
X Uric Acid	mg/dl	2.43	4.36	-4.86	-1.16
@ Urea	mg/dl	39.20	40.67	0.54	1.08
@ Creatinine	mg/dl	0.60	0.85	-1.91	-0.42
@ Bilirubin Total	mg/dl	0.60	0.45	1.33	0.85
@ Bilirubin Direct	mg/dl	0.32	0.22	0.90	0.81
@ Total Protein	g/dL	6.87	6.81	0.44	-0.50
@ Albumin	g/dL	4.00	4.20	-0.37	1.14
@ Calcium	mg/dl	9.40	8.82	1.30	0.88
@ Cholesterol	mg/dl	166.00	170.36	-0.51	-0.68
@ Triglycerides	mg/dl	116.00	137.36	-1.97	-0.47
@ HDL	mg/dl	23.00	31.25	-1.75	-0.33
@ LDL	mg/dl	94.96	99.01	-0.66	-0.33
@ ALT	U/L	18.80	21.41	-0.14	0.52
@ AST	U/L	30.90	41.16	-0.67	-0.63
@ ALP	U/L	60.84	70.28	-1.40	0.10

Instrument : ACCULYTE-3P

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Chloride	mmol/L	110.64	114.30	-0.26	-0.71
@ Sodium	mmol/L	162.07	162.40	-0.03	-0.09
@ Potassium	mmol/L	6.45	6.31	0.34	-0.19

Legend @ : Acceptable
 ! : 2.0 < z score < 3.0 - Warning
 X : z score ≥3 - Unacceptable
 # : Not Evaluated
 ⊙ : Delayed Result Entry
 * : Not considered for evaluation.

Total Parameters	18
Not Evaluated Parameters	0
Evaluated Parameters	18
Outlier Parameters (X)	1
EQAS Score Biochemistry	94.44 %



Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A
 Kohinoor Mall, Kirol Road, Kurla (W),
 Mumbai - 400070



Metropolis Healthcare Ltd., EQAS

KRSNAA DIAGNOSTICS,LTD,GURDASPUR (1039)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 240104 and Sample No 02

Report Date : 10/07/2024

Analyte	Instrument	Result Value	Standard Unit	Z-Score
! Bilirubin Direct	Diasys 200 Pro	0.20	mg/dl	2.1
Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : Diasys 200 Pro

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Glucose	mg/dl	85.50	71.18	0.56	0.75
@ Urea	mg/dl	23.50	26.57	-0.50	1.08
@ Creatinine	mg/dl	1.00	0.99	0.09	-0.42
@ Bilirubin Total	mg/dl	0.40	0.35	0.71	0.85
! Bilirubin Direct	mg/dl	0.20	0.14	2.10	0.81
@ Total Protein	g/dL	7.53	7.58	0.33	-0.50
@ Albumin	g/dL	4.10	4.19	0.04	1.14
@ Uric Acid	mg/dl	4.71	4.32	0.16	-1.16
@ Calcium	mg/dl	8.90	8.55	0.16	0.88
@ Cholesterol	mg/dl	170.00	162.79	0.05	-0.68
@ Triglycerides	mg/dl	135.00	135.91	-0.21	-0.47
@ HDL	mg/dl	32.50	29.24	0.10	-0.33
@ LDL	mg/dl	91.16	89.65	-0.32	-0.33
@ ALT	U/L	7.80	5.92	1.26	0.52
@ AST	U/L	14.66	13.94	0.00	-0.63
@ ALP	U/L	41.64	45.36	-0.88	0.10

Instrument : ACCULYTE-3P

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Chloride	mmol/L	100.12	105.34	-0.20	-0.71
@ Sodium	mmol/L	132.09	133.36	-0.15	-0.09
@ Potassium	mmol/L	5.19	5.13	0.06	-0.19

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning


X : z score ≥3 - Unacceptable

: Not Evaluated

⊙ : Delayed Result Entry

* : Not considered for evaluation.

Total Parameters	19
Not Evaluated Parameters	0
Evaluated Parameters	19
Outlier Parameters (X)	0
EQAS Score Biochemistry	100.00 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS

Unit No. 409-416.

Commercial Building - 1A

Kohinoor Mall, Kirol Road, Kurla (W),

Mumbai - 400070