

Analyte	Instrument	Result Value	Standard Unit	Z-Score
X Total Leucocytes (WBC) count	Horiba Yumizen H550	5330.00	1000/ μ L	6.12
! MCH (Mean Corpuscular Hb)	Horiba Yumizen H550	29.80	pg	2.87
! MCHC (Mean Corpuscular Hb Concentration)	Horiba Yumizen H550	32.00	g/dL	2.72
! RDW CV%	Horiba Yumizen H550	25.90	%	2.71
Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated Ⓞ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : Horiba Yumizen H550

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
X Total Leucocytes (WBC) count	/c.mm	5330.00	2655.00	6.12	0.27
@ Erythrocyte (RBC) Count	mill/cu.mm	2.33	2.46	-0.48	-0.70
@ Haemoglobin	g/dL	6.90	6.75	-0.42	-0.14
@ MCV (Mean Corpuscular Volume)	fL	92.80	106.18	-1.88	-1.62
! MCH (Mean Corpuscular Hb)	pg	29.80	27.27	2.87	1.41
! MCHC (Mean Corpuscular Hb Concentration)	g/dL	32.00	25.63	2.72	3.27
@ Platelet Count	1000/ µL	175.00	158.34	1.59	1.22
@ Hematocrit	%	21.70	26.11	-1.57	-1.67
! RDW CV%	%	25.90	18.58	2.71	3.37

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning


X : z score ≥3 - Unacceptable

: Not Evaluated

⌚ : Delayed Result Entry

* : Not considered for evaluation.

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	1
EQAS Score Haematology - CBC	88.89 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A
Kohinoor Mall, Kirol Road, Kurla (W),
Mumbai - 400070



Metropolis Healthcare Ltd., EQAS

KRSNAA DIAGNOSTICS,LTD,GURDASPUR (1039)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 240104 and Sample No 02

Report Date : 10/07/2024

Analyte	Instrument	Result Value	Standard Unit	Z-Score
! Total Leucocytes (WBC) count	Horiba Yumizen H550	3970.00	1000/ μ L	-2.87
! MCH (Mean Corpuscular Hb)	Horiba Yumizen H550	33.90	pg	2.22
X MCHC (Mean Corpuscular Hb Concentration)	Horiba Yumizen H550	35.20	g/dL	6.58
! Hematocrit	Horiba Yumizen H550	23.30	%	-2.06
X RDW CV%	Horiba Yumizen H550	30.50	%	6.45

Legend @ : z score \leq 2.0 - Acceptable
! : 2.0 < z score < 3.0 - Warning
X : z score \geq 3 - Unacceptable
: Not Evaluated
⊖ : Delayed Result Entry
* : Not considered for evaluation.

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : Horiba Yumizen H550

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
X MCHC (Mean Corpuscular Hb Concentration)	g/dL	35.20	24.80	6.58	3.27
X RDW CV%	%	30.50	18.04	6.45	3.37
@ Erythrocyte (RBC) Count	mill/cu.mm	2.42	2.73	-0.96	-0.70
! Total Leucocytes (WBC) count	/c.mm	3970.00	7787.78	-2.87	0.27
@ Haemoglobin	g/dL	8.20	8.06	-0.36	-0.14
@ MCV (Mean Corpuscular Volume)	fL	96.30	117.61	-1.93	-1.62
! MCH (Mean Corpuscular Hb)	pg	33.90	29.58	2.22	1.41
@ Platelet Count	1000/ µL	289.00	271.16	1.07	1.22
! Hematocrit	%	23.30	31.88	-2.06	-1.67

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning


X : z score ≥3 - Unacceptable

: Not Evaluated

⌚ : Delayed Result Entry

* : Not considered for evaluation.

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	2
EQAS Score Haematology - CBC	77.78 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS

Unit No. 409-416.

Commercial Building - 1A

Kohinoor Mall, Kirol Road, Kurla (W),

Mumbai - 400070



Metropolis Healthcare Ltd., EQAS

KRSNAA DIAGNOSTICS,LTD,GURDASPUR (1039)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 240104 and Sample No 03

Report Date : 10/07/2024

Analyte	Instrument	Result Value	Standard Unit
No Outlier			
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ Delayed Result Entry			

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____


Dated: _____

Instrument : -

Analyte	Standard Unit	Result Value	Accepted Value
@ Diagnosis (Slide)	-	CLL/Chronic lymphoproliferative disorder	CLL/Chronic lymphoproliferative disorder

Legend @ : Acceptable
 @ : Acceptable
 X : Unacceptable
 # : Not Evaluated
 ⌚ : Delayed Result Entry

Total Parameters	1
Not Evaluated Parameters	0
Evaluated Parameters	1
Outlier Parameters (X)	0
EQAS Score Haematology- Peripheral smear diagnosis	100.00 %



Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A
 Kohinoor Mall, Kiroi Road, Kurla (W),
 Mumbai - 400070



Metropolis Healthcare Ltd., EQAS

KRSNAA DIAGNOSTICS,LTD,GURDASPUR (1039)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 240104 and Sample No 03

Report Date : 10/07/2024

Analyte	Instrument	Result Value	Standard Unit
No Outlier			
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ Delayed Result Entry			

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____


Dated: _____

Instrument : -

Analyte	Standard Unit	Result Value	Accepted Value
@ Neutrophil.	-	<40%	<40%
@ Lymphocyte.	-	>40%	>40%
@ Eosinophil.	-	<6%	<6%
@ Monocyte.	-	1-10%	1-10%
@ Basophil.	-	<1%	<1%
@ Metamyelocyte.	-	Absent	Absent
@ Promyelocyte.	-	Absent	Absent
@ Myelocyte.	-	Absent	Absent

Legend @ : Acceptable
 @ : Acceptable
 X : Unacceptable
 # : Not Evaluated
 ⊙ : Delayed Result Entry

Total Parameters	8
Not Evaluated Parameters	0
Evaluated Parameters	8
Outlier Parameters (X)	0
EQAS Score Haematology-PS Slide Count Qualitative	100.00 %



Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A
 Kohinoor Mall, Kirol Road, Kurla (W),
 Mumbai - 400070