



Metropolis Healthcare Ltd., EQAS

KRSNAA DIAGNOSTICS,LTD,GURDASPUR (1039)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 240104 and Sample No 01

Report Date : 10/07/2024

Analyte	Instrument	Result Value	Standard Unit
X Protein (Urine)	Manual	1+	-
X Nitrite (urine)	Manual	Absent	-

Legend @ : Acceptable
X : Unacceptable
: Not Evaluated
⌚ Delayed Result Entry

Problem Classification: _____


Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : Manual

Analyte	Standard Unit	Result Value	Accepted Value
X Protein (Urine)	-	1+	2+
X Nitrite (urine)	-	Absent	Pre
@ pH	-	5	5
@ Specific Gravity	-	1.025	1.025
@ Glucose (urine)	-	3+	3+
@ Ketone (urine)	-	Negative	Negative
@ Blood (urine)	-	2+	2+
@ Bilirubin (urine)	-	Negative	Negative
@ Urobilinogen (urine)	-	Normal	Normal
Legend @ : Acceptable		Total Parameters	9
@ : Acceptable		Not Evaluated Parameters	0
X : Unacceptable		Evaluated Parameters	9
# : Not Evaluated		Outlier Parameters (X)	2
⌚ : Delayed Result Entry		EQAS Score Clinical Path - Urine Routine Chemistry	77.78 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A
Kohinoor Mall, Kirol Road, Kurla (W),
Mumbai - 400070