



Metropolis Healthcare Ltd., EQAS

KRSNAA DIAGNOSTICS,LTD,GURDASPUR (1039)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 240104 and Sample No 01

Report Date : 10/07/2024

Analyte	Instrument	Result Value	Standard Unit
X Widal test - Typhi O antigen (TO)	Manual	1:80	-
X Widal test - Typhi H antigen (TH)	Manual	1:160	-
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry			

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : Manual


Analyte	Standard Unit	Result Value	Accepted Value
X Widal test - Typhi O antigen (TO)	-	1:80	No Agglutination
X Widal test - Typhi H antigen (TH)	-	1:160	No Agglutination
@ RPR (Syphilis Serology)	-	Non-reactive	Non-reactive
@ Dengue NS 1	RU/ml	Negative	Negative
@ Widal test - ParaTyphi A-H antigen (AH)	-	No Agglutination	No Agglutination
@ Widal test - ParaTyphi B-H antigen (BH)	-	No Agglutination	No Agglutination
@ Dengue IgM	RU/ml	Negative	Negative
@ Dengue IgG	RU/ml	Negative	Negative

Instrument : J Mitra

Analyte	Standard Unit	Result Value	Accepted Value
@ HBsAg	IU/mL	Non-reactive	Non-reactive
@ HCV	IU/mL	Non-reactive	Non-reactive
@ HIV Ab	S/CO	Non-reactive	Non-reactive

Legend @ : Acceptable
 @ : Acceptable
 X : Unacceptable
 # : Not Evaluated
 ⌚ : Delayed Result Entry

Total Parameters	11
Not Evaluated Parameters	0
Evaluated Parameters	11
Outlier Parameters (X)	2
EQAS Score Immunoassay Extended	81.82 %



Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A
 Kohinoor Mall, Kirol Road, Kurla (W),
 Mumbai - 400070



Metropolis Healthcare Ltd., EQAS

KRSNAA DIAGNOSTICS,LTD,GURDASPUR (1039)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 240104 and Sample No 02

Report Date : 10/07/2024

Analyte	Instrument	Result Value	Standard Unit
X Widal test - Typhi O antigen (TO)	Manual	1:80	-
X Widal test - Typhi H antigen (TH)	Manual	1:80	-
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ Delayed Result Entry			

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : Manual


Analyte	Standard Unit	Result Value	Accepted Value
X Widal test - Typhi O antigen (TO)	-	1:80	No Agglutination
X Widal test - Typhi H antigen (TH)	-	1:80	No Agglutination
@ RPR (Syphilis Serology)	-	Non-reactive	Non-reactive
@ Dengue NS 1	RU/ml	Negative	Negative
@ Widal test - ParaTyphi A-H antigen (AH)	-	No Agglutination	No Agglutination
@ Widal test - ParaTyphi B-H antigen (BH)	-	No Agglutination	No Agglutination
@ Dengue IgM	RU/ml	Negative	Negative
@ Dengue IgG	RU/ml	Negative	Negative

Instrument : J Mitra

Analyte	Standard Unit	Result Value	Accepted Value
@ HBsAg	IU/mL	Reactive	Reactive
@ HCV	IU/mL	Non-reactive	Non-reactive
@ HIV Ab	S/CO	HIV-1/HIV-2 Ab Reactive	HIV-1/HIV-2 Ab Reactive

Legend @ : Acceptable
 @ : Acceptable
 X : Unacceptable
 # : Not Evaluated
 ⌚ : Delayed Result Entry

Total Parameters	11
Not Evaluated Parameters	0
Evaluated Parameters	11
Outlier Parameters (X)	2
EQAS Score Immunoassay Extended	81.82 %



Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A
 Kohinoor Mall, Kirol Road, Kurla (W),
 Mumbai - 400070



Metropolis Healthcare Ltd., EQAS

KRSNAA DIAGNOSTICS,LTD,GURDASPUR (1039)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 240104 and Sample No 03

Report Date : 10/07/2024

Analyte	Instrument	Result Value	Standard Unit
# HIV Ab	J Mitra	Non-reactive	-
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry			

Problem Classification:

Corrective Action:

Reviewed by:


Dated:

Instrument : J Mitra

Analyte	Standard Unit	Result Value	Accepted Value
@ HBsAg	IU/mL	Reactive	Reactive
@ HCV	IU/mL	Reactive	Reactive
# HIV Ab	S/CO	Non-reactive	

Legend @ : Acceptable
 @ : Acceptable
 X : Unacceptable
 # : Not Evaluated
 ⌚ : Delayed Result Entry

Total Parameters	3
Not Evaluated Parameters	1
Evaluated Parameters	2
Outlier Parameters (X)	0
EQAS Score Immunoassay Extended	100.00 %



Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A
 Kohinoor Mall, Kiroi Road, Kurla (W),
 Mumbai - 400070