

Date: 18-09-2023

Certificate of Calibration

Name & Address of Customer : NEHA PATHOLOGY LAB
..... 391, RAMNAWAR, ROORKEE
STATE..... Uttarakhand PIN:..... 247667
Phone:..... 9027979664 E-Mail :..... C.maurya.83@gmail.com

Name of Instrument: **Medonic M20/ M32B/ M32M**

Type : Hematology Analyzer

Serial No : 120747

Calibration Date : 18-09-2023

Next Calibration Due : 18-09-2024

This is to certify that above said instrument has been validated by hardware calibration for Background count, Aspiration (Pump) & Temperature. Five main parameters (WBC,RBC,HGB,MCV,PLT) has been calibrated according to the procedures provided by Boule Medical AB, Sweden

This calibration is carried out by using Standard Operating Procedures (S.O.P.) provided by Boule Medical AB, Sweden.

This instrument conform to IVD & EU directives of use.

Calibration carried out on site by:

Signature

Name of Service Engineer : Mukesh

Instrument Service Division

Q-LINE BIOTECH Pvt. Ltd