

Date: _____

Certificate of Calibration

Name & Address of Customer : PATHLINE DIAGNOSTIC LAB
Near Dr. Shushila Tiwari Medical College
Rampur Road, Haldwani
STATE Uttarakhand PIN: 263139
Phone: 8077416413 E-Mail: praveengirish@gmail.com

Name of Instrument: **Medonic M20/ M32B/ M32M**

Type : Hematology Analyzer

Serial No : 124165

Calibration Date : 17.02.2024

Next Calibration Due : 16.02.2025

This is to certify that above said instrument has been validated by hardware calibration for Background count, Aspiration (Pump) & Temperature. Five main parameters (WBC,RBC,HGB,MCV,PLT) has been calibrated according to the procedures provided by Boule Medical AB, Sweden

This calibration is carried out by using Standard Operating Procedures (S.O.P.) provided by Boule Medical AB, Sweden.

This instrument conform to IVD & EU directives of use.

Calibration carried out on site by:

Signature Mukesh
Name of Service Engineer : MUKESH KUMAR
Instrument Service Division
Q-LINE BIOTECH Pvt. Ltd