



Metropolis Healthcare Ltd., EQAS

KRSNAA DIAGNOSTICS LTD ROHRU (1240)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 240104 and Sample No 01

Report Date : 10/07/2024

Analyte	Instrument	Result Value	Standard Unit
X Glucose (urine)	Aspen U500	Negative	-

Legend @ : Acceptable
X : Unacceptable
: Not Evaluated
ⓐ : Delayed Result Entry

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : Aspen U500

Analyte	Standard Unit	Result Value	Accepted Value
X Glucose (urine)	-	Negative	+2,+3,+4
@ pH	-	6	5.5,5.6
@ Specific Gravity	-	1.020	1.020,1.025,1.030
@ Protein (Urine)	-	2+	+1,+2,+3
@ Ketone (urine)	-	Negative	+1,Neg,Trace
@ Blood (urine)	-	2+	+1,+2,+3

Instrument : -

Analyte	Standard Unit	Result Value	Accepted Value
@ Urine Pregnancy Test	-	Positive	Positive

- Legend @ : Acceptable
 @ : Acceptable
 X : Unacceptable
 # : Not Evaluated
 ⊙ : Delayed Result Entry

Total Parameters	
Not Evaluated Parameters	
Evaluated Parameters	
Outlier Parameters (X)	
EQAS Score Clinical Path - Urine Routine Chemistry	85.71 %

Dr Puneet Kumar Nigam

Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A
 Kohinoor Mall, Kirol Road, Kurla (W),
 Mumbai - 400070



Metropolis Healthcare Ltd., EQAS

KRSNAA DIAGNOSTICS LTD ROHRU (1240)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 240104 and Sample No 01

Report Date : 10/07/2024

Analyte	Instrument	Result Value	Standard Unit
No Outlier			
Legend @ : Acceptable			
X : Unacceptable			
# : Not Evaluated			
⊙ Delayed Result Entry			

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____



Metropolis Healthcare Ltd., EQAS

KRSNAA DIAGNOSTICS LTD ROHRU (1240)

Outlier And Analyte Summary Report

Result Details For Cycle No 240104 and Sample No 01

Report Date : 10/07/2024

Instrument : -

Analyte	Standard Unit	Result Value	Accepted Value
@ Crystal	-	Triple phosphate crystal	Triple phosphate crystal

Legend @ : Acceptable
@ : Acceptable
X : Unacceptable
: Not Evaluated
Ⓢ : Delayed Result Entry

Total Parameters	1
Not Evaluated Parameters	0
Evaluated Parameters	1
Outlier Parameters (X)	0
EQAS Score Clinical Path - Urine Sediment Analysis Image	100.00 %

Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A
Kohinoor Mall, Kiroi Road, Kurla (W),
Mumbai - 400070

CORRECTIVE ACTION & PREVENTIVE ACTION (CAPA)

Name of the Department:

Sr. No	Date	Non Conformity Observed	Root Cause Analysis	Corrective Action & Preventive Action	Date of Closure	Signature
1	9-8-24	Glucose urine	Random error	Spelid test done	9-8-24	[Signature]

Issued by: Vinod Lonkar [Signature]
 Approved by: Dr Amol Patil
 Doc No: CAPA-69

Issued on: 01.03.2024

Version No: _____
 m. id No: _____



SPLIT TESTING RECORD

Sr.No.	Date	Analyte Name	Barcode ID	Initial Result	Lab BRI	Initial test done by	Split test result	Split test done by	Acceptability criteria applied	Acceptability Range	Split test result acceptable (Yes / No)	Comment if not applicable
1	7 Aug 24	Albumin urine	2762	Absent	Absent	Saite	Absent	Jignay	N/A	-	Yes	-
2												
3												
4												
5												
6												

Entered by

Reviewed by

Approved by

Issued by : Vinod Lonkar <i>Vinod</i>	Issue Date : 01.09.2023
Approved by : Dr. Manish Karekar <i>Manish</i>	Amend No. : -
Doc No. GEN-14	Amend Date :
Version No : 1	