

HORIBA
MEDICAL

HORIBA India Private Limited

246, Okhla Industrial Estate Phase-III,
New Delhi 110020, India
Tel : +91 (11) 4646 5000 / 4669 5001
Fax : +91 (11) 4646 5020
http : //www.horiba.com
CIN : U73100DL2006PTC153232

HIN/MED/2023-2024/1004119
25TH AUG 2023

CALIBRATION CERTIFICATE

This is to certify that the Hematology Analyzer Horiba YUMIZEN bearing serial number 909YAXH02588 installed at **Krsnaa Diagnostics Ltd. – Rampur Bushahr** as calibrated on 25TH AUG 2023.

Calibrator : ABX Minocal
Lot No. : CX486
Expiry Date : 05TH OCT 2023

The reports of Blank Cycle, Repeatability and Calibration Values were all found in acceptable range.

Next calibration cycle is due on 24TH AUG 2024



Shrish Dixit

(Head- Products & Marketing)
For **Horiba India Pvt. Ltd.**

246, Okhla Industrial Estate Phase-III,
New Delhi 110020, India
Tel : +91 (11) 4546 5000 / 46695001
Fax : +91 (11) 4646 5020
http : //www.horiba.com
CIN : U73100DL2006PTC153232