

CALIBRATION CERTIFICATE

Calibration of the **LIFOTRONICS-H9** (HbA1c Analyzer) has been carried out as per the recommendations of the principle company **SHENZHEN LIFOTRONIC TECHNOLOGY CO., LTD, China.**

Customer Name & Address: KRSNAA DIAGNOSTIC LTD. C/O MAHATMA GANDHI MEDICAL SERVICES COMPLEX, KHANERI, RAMPUR BUSHAHR, SHIMLA, HIMACHAL PRADESH-172001	Installation Date : 19.09.2022 Calibration Date : 18.09.2023 Calibration Due Date: 17.09.2024
--	--

Serial No:	IA2B00001636
-------------------	---------------------

Software Version Information:

System Version:	01.00.022-S
MCU1 Version:	01.00.02-H
MCU2 Version:	01.00.01-G

Monitoring Information:

Test Mode	HbA1c Mode
HP Pressure pump stepping Value	105
Column Temp	41
H9 Temp	40.4
N.P.	66
ADC 415	24342
ADC 500	17460
ADC	7022