



A A CALIBRATION PVT. LTD.

CALIBRATION CERTIFICATE

7.8F



CC - 2646

Certificate No	AACPL / 08243F	Field	Elec. Tech.	Group		Page No.	1 of 1
SRF No & Date	231108.1	DT - 08 / 11 / 2023		ULR No.		CC2646230000	08243F
Calibration Method :	Direct Method			Accuracy			
Company Name	M/s Krsnaa Diagnostic Ltd.			Calibration Date	08.11.2024		
Address	Mahatma Gandhi Medical Services Complex, Khaneri, Rampur, Himachal Pradesh			Recommended Date for Next Calib.	07.11.2023		
				Certificate Issue Dt.	10.11.2023		
				UUC Condition	Good		

Calibration Instrument Detail

SI	Instrument Name	Range	Least Count	Make / SI.No.	Party ID / Location
1	Dig. Timer (Centrifuge M/c)	0 - 99 min	1 min	Remi ZIFN-26902	KDHPRMMSSE29 Serology

Standard Equipments Used (Traceable to National Standard)

SI	Instrument Name	Make/SI No	Calibrated By	Cal Certificate No	Cal. Valid Upto
	Dig. Time Interval Calibrator	Voltagram / ----	VTC	CC360623000002157F	18/10/2024

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration performed At
NIST 960 - 12	Time Measurement	57 % RH	25.4 °C	Site

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 03, Section No : CPET - 02

Calibration Result

SI	Selected Range (In s)	Average Read. (Std.) (In s)	Average Read. (UUC) (In min)	Deviation (In s)	Uncertainty at approx 95% confidence Level and coverage factor $k = 2$ is \pm
1	600	599.82	10	0.18	34.64 s
2	900	899.79	15	0.21	34.67 s
3	1800	1799.4	30	0.6	34.67 s

Note

- The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- This certificate cannot be reproduced except in full without our prior permission in writing
- This certificate refers only to the particular item(s) submitted for calibration
- UUC - Unit under calibration

----- End of Certificate -----

Calibrated by



Approved by

B. Kumar
Bhoopesh Kumar
(Tech. Manager)

LAB :- Plot No. C-18, Sector-13, First Floor, "Vasundhara, Haat" Vasundhara, Ghaziabad - 201012
Ph. : 0120-4135529, Mob. : 9212472564
E-mail : aacplinfo@aacpl.com