



A A CALIBRATION PVT. LTD.

CALIBRATION CERTIFICATE

7.8F-01



CC - 2546

Certificate No	AACPL / 08235F	Field	Elec. Tech.	Group	Page No.	1 of 1
SRF No & Date	231108.1	DT - 08 / 11 / 2023		ULR No.	CC2646230000 08235F	
Calibration Method :	Direct Method			Accuracy		
Company Name	M/s Krsnaa Diagnostic Ltd.			Calibration Date	08.11.2024	
Address	Mahatma Gandhi Medical Services Complex, Khaneri, Rampur, Himachal Pradesh			Recommended Date for Next Calib.	07.11.2023	
				Certificate Issue Dt.	10.11.2023	
				UUC Condition	Good	

Calibration Instrument Detail

SI	Instrument Name	Range	Least Count	Make / SI.No.	Party ID / Location
1	Dig. Timer (Centrifuge M/c)	0 - 99 min	1 min	Remi ZIFN-26904	KDHPRMMSCC16 Clinical Chemistry

Standard Equipments Used (Traceable to National Standard)

Instrument Name	Make/SI No	Calibrated By	Cal Certificate No	Cal. Valid Upto
Dig. Time Interval Calibrator	Voltagram / -----	VTC	CC360623000002157F	18/10/2024

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration performed At
NIST 960 - 12	Time Measurement	57 % RH	26.2 °C	Site

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 03, Section No : CPET - 02


Calibration Result

SI	Selected Range (In s)	Average Read. (Std.) (In s)	Average Read. (UUC) (In min)	Deviation (In s)	Uncertainty at approx 95% confidence Level and coverage factor $k = 2$ is \pm
1	600	599.77	10	0.23	34.64 s
2	900	899.75	15	0.25	34.67 s
3	1800	1799.2	30	0.8	34.67 s


Note

- The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- This certificate cannot be reproduced except in full without our prior permission in writing
- This certificate refers only to the particular item(s) submitted for calibration
- UUC - Unit under calibration

----- End of Certificate -----


Calibrated by



Approved by

Bhoopesh Kumar
(Tech. Manager)

LAB :- Plot No. C-18, Sector-13, First Floor, "Vasundhara, Haat" Vasundhara, Ghaziabad - 201012
Ph. : 0120-4135529, Mob. : 9212172564, 9212172597
E-mail : aacplinfo2005@gmail.com, aacplinfo@rediffmail.com