



# AACPL CALIBRATION PVT. LTD.

## CALIBRATION CERTIFICATE

7.8F-01



CC - 2646

Certificate No	AACPL / 08257F	Field	Mech.	Group	Mass	Page No.	1 of 1
SRF No & Date	231108.3 DT - 08 / 11 / 2023			ULR No.	CC2646230000 08257F		
Calibration Method :	Comparison Method			Accuracy			
Company Name	M/s Krsnaa Diagnostic Ltd.			Calibration Date	08.11.2024		
Address	Mahatma Gandhi Medical Services Complex, Khaneri, Rampur, Himachal Pradesh			Recommended Date for Next Calib.	07.11.2023		
				Certificate Issue Dt.	10.11.2023		
				UUC Condition	Good		

### Calibration Instrument Detail

Sl	Instrument Name	Range	Least Count	Make / SI.No.	Party ID / Location
1	Micro Pipette	10 to 100 $\mu$ l	1 $\mu$ l	Dragon Lab YE22AX0053958	KDHPRMMSSE25 Serology

### Standard Equipments Used ( Traceable to National Standard )

Sl	Instrument Name	Make/SI No	Calibrated By	Cal Certificate No	Cal. Valid Upto
1	Weighing Balance	Radwag / 668683	Bluebox Techno	BBT/452/JAN/23	05/03/2024

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
ISO - 8655 - 6	Volume Measurement	54 %RH	23.1 $^{\circ}$ C	Laboratory

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 01C, Section No : CP - 03

### Calibration Result

S. No	Nominal Value on UUC (in $\mu$ l)	Average Volume at 27 $^{\circ}$ C (in $\mu$ l)	Deviation (in $\mu$ l)	Uncertainty at approx 95% confidence Level and coverage factor $k = 2$ is $\pm$ (in $\mu$ l)
1	10	10.01728	-0.01728	0.25
2	50	50.03153	-0.03153	0.25
3	100	100.05326	-0.05326	0.25

#### Note

- The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- This certificate cannot be reproduced except in full without our prior permission in writing
- This certificate refers only to the particular item(s) submitted for calibration
- UUC - Unit under calibration

----- End of Certificate -----

Calibrated by  
*Ashish*



Approved by  
*Aashid Chaurasia*  
Aashid Chaurasia  
(CEO)