



The Accurate Calibration & Medical Solution

CALIBRATION CERTIFICATE

Page 1 of 1

CALIBRATION CERTIFICATE OF MICROSCOPE

Certificate No :	TAC/11-23/KDL/MS-1002	Calibration Date :	08.11.2023
Job Identification No :	TAC/C/08/11/2023	Next Calibration Date :	07.11.2024
Ref. No :	SRF Dated - 06. 11. 2023	Issue Date :	10.11.2023

Customer Name	M/s Krsnaa Diagnostic Ltd.
Address	Mahatma Gandhi Medical Services Complex, Khaneri, Rampur, Himachal Pradesh 172001

DESCRIPTION OF THE EQUIPMENT TO BE CALIBRATED

Instrument Name	Micro Scope	Visual Inspection	Ok
Make	Glomax	I.D. No.	KDHPRMMSCP23
Model	-----	Location / Dept.	Clinical Pathology
Serial No.	-----		

CALIBRATION STANDARD USED

Sl.No.	Master Equipment	Manufacturer	Model No.	Serial No.	Date of Calibration	Valid Upto
1	Electrical Safety Analyzer	DATREND Systems Inc.	vPed - Mini	VPM18100028	14.10.2023	14.10.2024

ENVIRONMENTAL CONDITIONS	Temperature	22.0 °C ± 2.0 °C	Humidity	50 ± 10%RH
--------------------------	-------------	------------------	----------	------------

RESULT

VISUAL TEST

Sr.No.	Tests	Remarks
1	Power Cords, Cable Checking	OK
2	Mains Socket Checking	OK
3	Equipment Type (B, BF, CF)	CF
4	Equipment Class (I, II, III)	I


FUNCTIONAL TEST

Sr.No.	Parameters	Measured Value	Remarks
1	Voltage Between Live and Neutral (VLN)	231.8 V (220-240 V)	OK
2	Voltage Between Live and Earth (VLE)	230.2 V (220-240V)	OK
3	Voltage Between Neutral and Earth (VNE)	0.4 V (0-5V)	OK
4	Load Current (Ia)	0.4 A	OK
5	Enclosure Current	0.6 μA (0-10μA)	OK
6	Leaking Current (IL)	115 μA (0-300μA)	OK
7	Time for Testing (t)	15 (min)	OK

Electrical Safety Test Passed as per IEC Standard 60601.1

Calibration Passed

YES



NO



Remarks:

1. Certificates shall not be reproduced except the written approval of The Accurate Calibration & Medical Solution.
2. The Calibration result reported in this certificate are only valid at the time of Calibration and the stated condition of measurement.

Calibrated by



Approved by

Plot No. 72/G1, Ganesh Vihar Sirsi Road, Jaipur (Raj.) -302012

Mob.: 9672294758

E-mail : theaccuratecalibration@gmail.com