



**HORIBA India Private Ltd.**

246, Okhla Industrial  
Estate, Phase - III 110  
020 New Delhi, India

Tel: +91 (11) 4646 5000 / 4669 5001

Fax: +91 (11) 4669 5010 / 4646 5020

HIN/MED/2024-2025/1499

**24<sup>th</sup> April 2024.**

**CALIBRATION CERTIFICATE**

This is to certify that the Haematology Analyzer **ABX YUMIZEN H550** bearing serial number: **208YAXH03902** installed at **Krasna diagnostic center, Kapurthala** was Calibrated on **24<sup>th</sup> April 2024.**

Calibrator: ABX MINOCAL

Lot No.: CX493

Expiry Date: 05<sup>th</sup> May 2024.

The reports of Blank Cycle, Repeatability and Calibration Values were all found in acceptable range.

**Next calibration cycle is due on 23<sup>rd</sup> April 2025.**



**Shrish Dixit**

(Head- Products & Marketing)

For **Horiba India Pvt. Ltd.**