

Sensa Core Medical Instrumentation Pvt Ltd

Plot No.3,EPIP Zone,Pashamylaram, Medak Distric-502307. www.sensacore.com

Instrument Name: ST-200CL PLUS Na/K/Cl Electrolyte Analyzer

Date of Testing : 06/08/2024

Due Date:05/08/2025

Instrument Sr.No.: STCLP-0069

Calibration Slope :

S.NO	Na+	K+	Cl
Mean	52.58	58.56	41.76
Ranges	35-85 mV/decade	35-85 mV/decade	35-85 mV/decade
Status	OK	OK	Ok

Quality Control (I) :

S.NO	Na+	K+	Cl
Mean	126.8	2.20	83.2
Ranges	111 – 129 mmol/L	1.50 – 2.50 mmol/L	71-89 mmol/L
Status	OK	OK	OK

Quality Control (II) :

S.NO	Na+	K+	Cl
Mean	142.0	4.14	103.4
Ranges	135 – 149 mmol/L	3.60 – 4.60 mmol/L	93-107 mmol/L
Status	OK	OK	OK

Quality Control (III) :

S.NO	Na+	K+	Cl
Mean	162.4	6.04	122.0
Ranges	153 – 171 mmol/L	5.50 – 6.50 mmol/L	111-129 mmol/L
Status	OK	OK	OK

For, YUCCA DIAGNOSTICS.

Authorized Signatory with Stamp

Date: 06/08/2024



- Works : Plot No. A-49, Five Star M.I.D.C. Area, Kagal, Dist.Kollhapur-416236.
- Customer Care : (0231)2305087, Cell : 9370636578,9325657848,9823122822. WhatsApp No.: 8956437314.
- Email : yuccadiagnostics@gmail.com, Website : www.yucca.in