



# AACPL CALIBRATION PVT. LTD.

## CALIBRATION CERTIFICATE

7.8F-01



CC - 2646

Certificate No	AACPL / 03467F	Field	Elec. Tech.	Group	Page No.	1 of 1
SRF No & Date	240507.4 DT - 07 / 05 / 2024			ULR No.	CC2646240000 03467F	
Calibration Method :	Direct Method			Accuracy		
Company Name	M/s Krsnaa Diagnostics Ltd.			Calibration Date	07.05.2024	
Address	8QQ8+783, Regional Hospital Rd, Cinema Colony, Changer Sector, Kosrian Sector, Bilaspur, Himachal Pradesh			Recommended Date for Next Calib.	06.05.2025	
				Certificate Issue Dt.	13.05.2024	
				UUC Condition	Good	

### Calibration Instrument Detail

SI	Instrument Name	Range	Least Count	Make / SI.No.	Party ID / Location
1	Dig. Temp. Controller	2 to 8 °C	1 °C	Haier	KDHPBLRHBC16
				BS033CE8701T1MBT0969	Blood Collection (Refrigerator)

### Standard Equipments Used ( Traceable to National Standard )

SI	Instrument Name	Make/SI No	Calibrated By	Cal Certificate No	Cal. Valid Upto
1	Universal Calibrator	Radix / 112004873	AACPL	AACPL/02470F	01/04/2025

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
IS : 1248	Temp. Measurement	55 %RH	25.7 °C	Site

**Calibration Procedure :** As per Calibration Procedure Manual AACPL / CP / 03, Section No : CPET - 01

**Set Points :-** 2, 5 & 8 °C

Calibration Results				
SI. No.	Std. Reading on Master in °C	UUC Average Reading in °C	Deviation in °C	Uncertainty at approx 95% confidence Level and coverage factor $k = 2$ is $\pm$
1	2.1	3	0.9	1.62 °C
2	5.1	6	0.9	1.62 °C
3	8.1	10	1.9	1.62 °C

**Note**

- 1 The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- 2 This certificate cannot be reproduced except in full without our prior permission in writing
- 3 This certificate refers only to the particular item(s) submitted for calibration
- 4 UUC - Unit under calibration

----- End of Certificate -----

*[Signature]*  
Calibrated by



Approved by  
*[Signature]*  
Anand Chaurasia  
(CEO)