

## **A A CALIBRATION PVT. LTD.**

## **CALIBRATION CERTIFICATE**



CC - 2646

Certificate No	AACPL	03466F	Field	Elec. Tech.	Group	Page N		1 of 1
SRF No & Date	240507.4 DT - 07 / 05 / 2024		ULR No.	CC2646240000 03466F				
Calibration Method:	Direct Method		Accuracy					
Company Name	M/s Krsnaa Diagnostics Ltd.			Calibration		07.05		
Address	8QQ8+783, Regional Hospital Rd,			Recommend	ed Date for Home		.2025	
	Cinema Colony, Changer Sector,			Certificate	Issue Dt.	13.05	.2024	
			UUC Cond	lition	Good			

**Calibration Instrument Detail** 

SI	Instrument Name	Range	Least Count	Make / SI.No.	Party ID / Location
1	Dig. Temp. Controller	2 to 8 °C	1 °C	LG	KDHPBLRHHM15
				208PR22C104594	Hematology (Refrigerator)

Standard Equipments Used (Traceable to National Standard)

			to mational otalians /				
		Instrument Name Make/SI No		Calibrated By	Cal Certificate No	Cal. Valid Upto	
1	1	Universal Calibrator	Radix / 112004873	AACPL	AACPL/02470F	01/04/2025	

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
IS : 1248	Temp. Measurement	56 %RH	25.8 °C	Site

Calibration Procedure: As per Calibration Procedure Manual AACPL / CP / 03, Section No : CPET - 01

Set Points :- 2, 5 & 8 °C

Calibration Results							
SI. Std. Reading on Master UUC Average Reading Devi		Deviation	Uncertainty at approx 95% confidence				
No.	in °C	in °C	in °C	Level and coverage factor k =2 is ±			
1	2.1	2	-0.1	1.62 °C			
2	5.1	6	0.9	1.62 °C			
3	8.1	9	0.9	1.62 °C			

## Note

- 1 The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- 2 This certificate cannot be reproduced except in full without our prior permission in writing
- 3 This certificate refers only to the particular item(s) submitted for calibration
- 4 UUC Unit under calibration

----- End of Certificate -----





Anand Chaurasia (CEO)