



# A A CALIBRATION PVT. LTD.

## CALIBRATION CERTIFICATE



CC - 2646

Certificate No	AACPL / 04453F	Field	Mech.	Group	Volume	Page No.	1 of 1
SRF No & Date	240612.4	DT - 12 / 06 / 2024		ULR No.	CC2646240000 04453F		
Calibration Method :	Comparison Method			Accuracy			
Company Name	M/s Krsnaa Diagnostics Ltd.			Calibration Date	12.06.2024		
Address	Civil Hospital, Near Jyoti Chowk, Jalandhar			Recommended Date for Next Calib.	11.06.2025		
				Certificate Issue Dt.	13.06.2024		
				UUC Condition	Good		

### Calibration Instrument Detail

Sl	Instrument Name	Range	Least Count	Make / SI.No.	Party ID / Location
1	Micro Pipette	10 to 100 µl	1 µl	Dragon Lab YE223AX0089504	KDJLPBCCMP20 Clinical Chemistry

### Standard Equipments Used ( Traceable to National Standard )

Sl	Instrument Name	Make/SI No	Calibrated By	Cal Certificate No	Cal. Valid Upto
1	Weighing Balance	Radwag / 668683	Bluebox Techno	BBT/744/JAN/24	04/03/2025

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
ISO - 8655 - 6 - 2022	Volume Measurement	55 %RH	23.8 °C	Laboratory

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 01C, Section No : CP - 03

### Calibration Result

S. No	Nominal Value on UUC (in µl)	Average Volume at 27 °C (in µl)	Deviation (in µl)	Uncertainty at approx 95% confidence Level and coverage factor k = 2 is ± (in µl)
1	20	19.99854	0.00146	0.25
2	50	49.99823	0.00177	0.25
3	100	99.99789	0.00211	0.25

- Note
- 1 The calibration results reported in this calibration certificate are valid at the time of & under stated condition
  - 2 This certificate cannot be reproduced except in full without our prior permission in writing
  - 3 This certificate refers only to the particular item(s) submitted for calibration
  - 4 UUC - Unit under calibration

----- End of Certificate -----



*[Signature]*  
Calibrated by

Approved by

*[Signature]*