



## HORIBA India Private Limited

246, Okhla Industrial Estate Phase-III,  
New Delhi 110020, India  
Tel : +91 (11) 4646 5000 / 4669 5001  
Fax : +91 (11) 4646 5020  
http : //www.horiba.com  
CIN : U73100DL2006PTC153232

HIN/MED/2024-2025 /2017  
09<sup>th</sup> AUG 2024

## **CALIBRATION CERTIFICATE**

This is to certify that the Hematology Analyzer Horiba ABX MES60 bearing serial number 209ESOH20285 installed at **Krsnaa Diagnostics Ltd. - Bilaspur** as calibrated on 09<sup>TH</sup> AUG 2024.

Calibrator : ABX Minocal  
Lot No. : CX497  
Expiry Date : 05<sup>th</sup> Sep 2024

The reports of Blank Cycle, Repeatability and Calibration Values were all found in acceptable range.

Next calibration cycle is due on 08<sup>TH</sup> AUG 2025



Shrish Dixit

(Head- Products & Marketing)  
For **Horiba India Pvt. Ltd.**

246, Okhla Industrial Estate Phase-  
III, New Delhi 110020, India  
Tel : +91 (11) 4546 5000 / 46695001  
Fax : +91 (11) 4646 5020  
http : //www.horiba.com  
CIN : U73100DL2006PTC1M232

