

HORIBA

Medical

HORIBA India Private Limited

246, Okhla Industrial Estate Phase-III,

New Delhi 110020, India

Tel : +91 (11) 4646 5000 / 4669 5001

Fax : +91 (11) 4646 5020

http : //www.horiba.com

CIN : U73100DL2006PTC153232

HIN/MED/2024-2025/ 1811

26th JUN 2024


CALIBRATION CERTIFICATE

This is to certify that the Hematology Analyzer Horiba YUMIZEN 550 bearing serial number 902YAXH02142 installed at **Krsnaa Diagnostics Ltd. – Kullu** as calibrated on 26TH June 2024.

Calibrator : ABX Minocal
Lot No. : CX495
Expiry Date : 05th JULY 2024

The reports of Blank Cycle, Repeatability and Calibration Values were all found in acceptable range.

Next calibration cycle is due on 25TH JUNE 2025.

A circular stamp with the text "HORIBA INDIA PVT. LTD." around the perimeter. Overlaid on the stamp is a handwritten signature in black ink.

Shrish Dixit

(Head- Products & Marketing)
For **Horiba India Pvt. Ltd.**

246, Okhla Industrial Estate Phase-

III, New Delhi 110020, India

Tel : +91 (11) 4546 5000 / 46695001

Fax : +91 (11) 4646 5020

http : //www.horiba.com

CIN : U73100DL2006PTC1M232