



The Accurate Calibration & Medical Solution

CALIBRATION CERTIFICATE

Page 1 of 1

CALIBRATION CERTIFICATE OF MICROSCOPE

Certificate No :	TAC/06-24/KDL/MS-1001	Calibration Date :	11.06.2024
Job Identification No :	TAC/C/11/06/2024	Next Calibration Date :	10.06.2025
Ref. No :	SRF Dated - 09. 06. 2024	Issue Date :	13.06.2024

Customer Name	M/S KRSNAA DIAGNOSTICS LTD.
Address	CIVIL HOSPITAL NEAR JYOTI CHOWK JALANDHAR - 144001

DESCRIPTION OF THE EQUIPMENT TO BE CALIBRATED

Instrument Name	MICRO SCOPE		
Make	LABOMED	Visual Inspection	Ok
Model	LX300	I.D. No.	KDJLPBHMS16
Serial No.	210630017	Location / Dept.	HAEMATOLOGY

CALIBRATION STANDARD USED

Sl.No.	Master Equipment	Manufacturer	Model No.	Serial No.	Date of Calibration	Valid Upto
1	Electrical Safety Analyzer	DATREND Systems Inc.	vPed - Mini	VPM18100028	14.10.2023	14.10.2024

ENVIRONMENTAL CONDITIONS	Temperature	22.0 °C ± 2.0 °C	Humidity	50 ± 10%RH
--------------------------	-------------	------------------	----------	------------

RESULT

VISUAL TEST

Sr.No.	Tests	Remarks
1	Power Cords,Cable Checking	OK
2	Mains Socket Checking	OK
3	Equipment Type (B,BE,CF)	CF
4	Equipment Class (I, II, III)	I


FUNCTIONAL TEST

Sr.No.	Parameters	Measured Value	Remarks
1	Voltage Between Live and Neutral (VLN)	231.9 V (220-240 V)	OK
2	Voltage Between Live and Earth (VLE)	229.5 V (220-240V)	OK
3	Voltage Between Neutral and Earth (VNE)	0.4 V (0-5V)	OK
4	Load Current (Ia)	0.3 A	OK
5	Enclosure Current	0.4 µA (0-10µA)	OK
6	Leaking Current (IL)	106 µA (0-300µA)	OK
7	Time for Testing (t)	15 (min)	OK

Electrical Safety Test Passed as per IEC Standard 60601.1

Calibration Passed

YES



NO




Remarks:

- 1.Certificates shall not be reproduced except the written approval of The Accurate Calibration & Medical Solution.
- 2.The Calibration result reported in this certificate are only valid at the time of Calibration and the stated condition of measurement.

Calibrated by

**** END CERTIFICATE ****



Approved by