

Diagnosics India Private Limited

Industrial Area,

DiaSys

229, FIA, Patparganj

Ground and first Floor,

New Delhi -110092



CALIBRATION CERTIFICATE

CALIBRATION PROTOCOL

The purpose of this calibration protocol is to define the qualifications and the acceptance standard in order to verify the normal operation and function of the Diasys SYS1000 Auto chemistry analyzer in the laboratory. Trained knowledgeable personnel from Diasys Diagnostics India Pvt Ltd along with the department personnel will perform and review analyser calibration protocol as mentioned by the manufacturer. The satisfactorily outcome of this procedures assures that the system functions according to the parameters.

EQUIPMENT INFORMATION: -

Instrument NameSYS1000 : AUTO CHEMISTRY ANALYZER

Model/TypeSYS1000 / : Bepnch Type-Random Bio Chemistry Analyzer

Serial NoA042307009 :

Calibration Done On10-08- : 2024

Next Calibration Due On10- : 08-2025

: Laboratory/HospitalKRSNAA DIAGNOSTIC LTD (Inside
Regional hospital Kullu (H.P)

Supported ByDiasys : Diagnostics India Pvt Ltd

Diagnosics India Private Limited

Industrial Area,

CIN

Registered Office

Plot No. A.821, T.T.C. Area, DiaSys Diagnostics India Private Limited

IDC, Bhubane, Navi Mumbai -400710 The Specialists for diagnostic system solutions

Maharashtra, India Founding and Automated Laboratory systems Email • info@diasys.in Tel. No.

Toll Free 18001201447

Website: MH.diasys-diagnostics.com

Diagnostics India Private Limited


Industrial Area,

DiaSys Diagnostics India Private Limited

229, FIA, Patparganj Industrial Area,

Ground and First Floor,

New Delhi -110092



CALIBRATION AND MAINTAINANCE PROCEDURES

MAINTAINANCE:

Instrument was checked for cleanliness.

Cuvettes were manually cleaned and also performed rinsing cuvette operation in maintenance.

Wash unit working was checked whether it is dispensing and aspirating alkaline detergent and water in the perfect sequence manually

- Cleaned sample, reagents and mixing probes manually with alcohol and then with distilled water, also performed sample probe rinsing in the maintenance.
- Performed sample, reagents and mixer probes horizontal and vertical check-up.
- Auto effluent pipeline rinsing was performed using magnetic valve detergent.
- Syringe exhaust was performed to remove any air bubbles present in the syringes.
- Instrument mechanism operation check was done by 20 times and found instrument satisfied operation conditions.

CALIBRATION:

- Input supply and lamp voltage for the machine was found adequate using the multimeter.
- Lamp intensity and cuvette check values are measured in the maintenance screen and reports are attached
- Incubation temperature readings are indicated as in the software
- Reagent compartments cooling temperature is indicated in LED display behind the machine.
- Voltages accross different testing points were checked.

Diagnostics India Private Limited

Industrial Area,

CIN : U74999MH2013PTC243371

eglttrtd once

Email :

DiaSys Diagnostics India Private Limited

TeLNm :.stu-

ü331

IIDC, b%ape, Na'd

-400710 Ttb

for &gnostic

ToUFree : ahataashtra, Ma

Sys

DiaSys

229, FIA, Patparganj Ground and First Floor,
New Delhi -110092

India

This is to certify that this Analyzer has been inspected and calibrated for following parameters

TEST PARAMETER	TARGET VALUE & RANGE	OBTAINED VALUE
INPUT VOLTAGE	230-240V AC	234V AC
LAMP INTENSITY CHECK	8000- 18000 for all wavelengths	Refer attachment
CUVETTE CHECK	<+/- 800 for all wavelengths	Refer attachment
INCUBATOR TEMP	37 degree +/- 0.3	37.0 degree (Software)
REAGENT COMPARTMENT COOLING CHECK	4to 10 degrees centigrade	5.2 degree
DETECTOR PERFORMANCE CHECK	Voltage b/w 3.5 to 4.5V	3.7 to 4.2V
12V LAMP SUPPLY	12 +/- 0.3V	12.18V
5V SUPPLY	5 +/- 0.3V	5.05V
24V SUPPLY	24 +/- 0.3V	23.98V

The results are obtained as per specifications & tolerance ranges. Calibrations of routine tests were also done & the results of controls & sample repeatability found satisfactory.

Report Sign Off:

Calibration Done By: Gaurav Dhand

Designation: Sr. Area Service Manager

Date & Sign: 10/12/2022

Gaurav Dhand



Diagnostics India Private Limited

Industrial Area,

Registered Office

No. A-821, T.T.C. Industrial Area,

IDC, Mahap, Na'a MINTE- "710

ima

DiaSys Diagnostics India private Limited

The Specialists for diagnostic system solutions

P&I-d we lashg and aulwrved LaboratOff systems

Email : dia@dia.com
Tel. No. : 491-022-71068331

Toll Free : 1800 120 1447
www.diasys-diagnostics.com

U7.BEMH20i3PTC2<S371

CIN:

Maharashtra,