

		CLINICAL BIO CHEMISTRY
	EQAS- ROOT CAUSE ANALYSIS FORM	

EQAS/ILC Programme or Provider Name: CMC ERAS					
Month & Year: April 2024		Cycle No: ✓		Sample No.: ✓	
1. Date sample received		Received by (sign):	2. Dates(s) Analysis performed	13.4.24	Done By (sign): M. [Signature]
3. Date results submitted	13.4.24	Submitted by (sign):	4. Date results/feedback received	8.5.24	Reviewed By (sign): M. [Signature]

DETAILS OF UNACCEPTABLE RESULT :

S.No	Analyte Name	Reported result	Result received (Scores)	Acceptable limits
1.	Cholesterol	55.7 ^{2L}	-3.41	$ z < 3.0$

EVALUATION OF POSSIBLE SOURCES OF ERROR

1. Clerical error	Yes	No	N A
Was the correct result / value transcribed from the instrument read out to workbook	✓		
Was the correct result / value transcribed from the workbook to EQAS report	✓		
Do the units of measure match between results form and instrument / your lab's unit	✓		
Is the decimal place correct / misplaced	✓		
Does the result reported by you in the result form match with that in the evaluation report	✓		
2. Procedural errors	Yes	No	N A
Was the written / usual procedure followed	✓		
Were reagents prepared according to procedure	✓		
Were the reagents within the open vial stability period (refer kit in use log)	✓		
Were Internal Quality control results acceptable on day of testing EQAS specimens for level I & II	✓		

		CLINICAL BIO CHEMISTRY
	EQAS- ROOT CAUSE ANALYSIS FORM	

3. Analytical Errors:	Yes	No	N A
Was the calibration for the analyte in question done?	✓		
Was calibration status live or expired?	✓		
Was the most recent calibration result acceptable and within established stability limits at the time of EQAS / PT testing?	✓		
Was calibration done after the lot of reagent in use was put into use?	✓		
If there was a major equipment breakdown / maintenance after last EQAS Was recalibration done	✓		
Are the assay parameters in the equipment are as per product insert of analyte ?	✓		
Was the result unacceptable for the same analyte in the previous 2 cycles of EQAS ?		✓	
Are results for the analyte evenly distributed in previous EQAS cycle	✓		
Is there a trend or bias on review of previous EQAS results?		✓	
Do EQAS results show unacceptable results for same range of value (high, normal, low)		✓	
Were the intended results within the measuring range of instrument /kit?	✓		
Was any dilution done during analysis?		✓	
If dilution was done were results multiplied by dilution factor and sent to EQAS provider?		✓	
4. Instrument Maintenance	Yes	No	N A
Were there any equipment maintenance / breakdown just after EQAS /PT specimen analysis (next 1 week period) that could have affected results? (Review Equipment breakdown log)		✓	
Were there any equipment maintenance / change of critical parts just prior to EQAS /PT specimen processing? (Review Equipment breakdown log)		✓	
Were new reagents used / reagent lot changed just prior or after EQAS /PT specimen processing that could have caused problems because of QC results and calibration expiry ?		✓	
5. PT specimen	Yes	No	N A
Were EQAS /PT samples reconstituted correctly according instructions of EQAS provider? Volume of DW, use of glass pipette, swirling, complete dissolution, storage	✓		
Was the water quality good and checked recently	✓		
Were any special instructions provided by the kit instructions performed as indicated?	✓		
Were the correct tests performed on the correct vial of PT specimen? (if more than 1 PT specimen was provided)	✓		
Were your results graded with the appropriate peer group based on the method reported on the result form sent by you ?	✓		

		CLINICAL BIO CHEMISTRY
	EQAS- ROOT CAUSE ANALYSIS FORM	

If **answer** to any of the **above Questions** was '**NO**' record the **corrective action taken** here:

Random error - may be due to pipetting,
 training given for proper reconstitution of
 EQAS sample to technicians.

Investigated By	1 st Review By	2 nd Review By
S. Kanitha	G. Kanitha	M. E. Thi (M. M. Gomathi)
Senior Lab technician	Quality Manager	Lab Director
Date: 10.5.24	Date: 10.5.24	Date: 10.5.24