





TATA 1mg Labs	
TATA 1mg Technologies Private Limited	
Form Name	THERMOMETER/DIGITAL SENSOR CALIBRATION
Form No.	Gen/FR/112
Issue Date & Version No.	01-Oct-2021 V1

Calibration Done Date: 30-07-2023	Next Due Date: 29-07-2024
Report No:	
Acceptable Limit for Variation in Temperature : $\pm 1^{\circ}\text{C}$	
Master's Sr. No. / ID Used for Calibration:	GT-01
Traceable to NABL Accredited Lab Report No. / Date: 23100002402 / 6.2.23	
Department:	LAB
Thermometer ID:	TH/LAB-03
Instrument Name:	Refrigerator
Asset ID:	TATA1MG/ANGUL/LAB/REF-03

Temperature Calibration

Sr.No.	Set Temperature for the Calibrated Value	Actual Value as per the Master Report	Average of 5 Measured Temperature of Reference Calibration Thermometer	Standard Deviation	Error	Expanded Uncertainty in $^{\circ}\text{C}(\pm)$
1	2	2.10	1.98	0.13	-0.12	0.61
2	5	5.11	5.04	0.11	-0.07	0.60
3	8	8.12	8.00	0.10	-0.12	0.60

Calibration Status: OK	
Performed By: Amit Kumar	Reviewed by: Mahipal Singh
 Sign	 Sign
30/07/23 Date / Time	01/08/2023 Date / Time

Controlled Copy