

PARTICIPANT FINAL ASSESSMENT REPORT (SUMMARY)

PT SCHEME:HEMATOLOGY PLUS

Cycle No	C6
Ref.No.	NEUQAP644
Sample ID	NEUQAP HEMOS BLOOD/2024/7/C6/S4

Report Date : 8/7/2024

Sample : JUL 2024

Peer group

	Parameters	Labs (n)	Lab Value	Unit	Assigned value	SDPA	Um	Z-Score	Remarks
17	HEMOGLOBIN (Automated Cell counter)	13	13.10	g/dL	13.54	* 0.26	0.084	1.73	Satisfactory
18	PLATELET COUNT (Automated Cell counter)	14	173.00	10 ³ /c umm	199.77	* 22.19	7.267	1.21	Satisfactory
19	TOTAL COUNT (Automated Cell counter)	13	10.14	10 ³ /c umm	7.71	1.78		1.36	Satisfactory
20	RBC (Automated Cell counter)	26	4.25	10 ⁶ /c umm	4.38	0.09	0.023	1.44	Satisfactory
21	MCV (Automated Calculator)	2	97.70	fl	94.05	5.16		0.71	Good
22	MCH (Automated Calculator)	2	30.80	pg	31.05	0.35		0.71	Good
23	MCHC (Automated Calculator)	2	31.50	g/dL	33.05	2.19		0.71	Good
24	Neutrophils (Automated Cell counter)	27	56.00	%	50.79	8.27	1.990	0.63	Good
25	Lymphocyte (Automated Cell counter)	28	36.00	%	36.11	6.70	1.583	0.02	Good
26	Monocyte (Automated Cell counter)	27	5.00	%	5.81	1.75	0.420	0.46	Good
27	Eosinophils (Automated Cell counter)	28	3.00	%	4.99	2.37	0.560	0.84	Good
28	Basophils (Others)	0	Not Recvd.	%					N/A

Authorised Signatory



Dr. Sujay Prasad
Technical Manager and Program coordinator

* denotes adjusted SDPA as per ISO 13528 standard
SDPA - Standard Deviation for Proficiency Assessment

Verified By



Dr. Glen
Page 3 of 33

This report is for use only by the intended participant.



An external quality assurance program from
NEUBERG ANAND ACADEMY OF LABORATORY MEDICINE PVT.LTD.

Anand Tower, 54, Bowring Hospital Road,
Shivajinagar, Bangalore - 560001
+91-9902277355 · qap@naalm.com · www.naalm.com
CIN: U85300KA2018PTC115147