

HORIBA
Explore the future

HORIBA India Private Ltd.

246, Okhla Industrial

Estate, Phase - III 110

020 New Delhi, India

Tel: +91 (11) 4646 5000 / 4669 5001

Fax: +91 (11) 4669 5010 / 4646 5020

HIN/MED/2024-2025/120
19th June 2024.

CALIBRATION CERTIFICATE

This is to certify that the Hematology Analyzer ABX MICROS ES 60 bearing serial number: 20SESOH20250 installed KRSNA DIAGNOSTIC LAB, Himachal Pradesh was Calibrated on 19th June 2024.

Calibrator: ABX MINOCAL

Lot No.: CX495

Expiry Date: 05th July 2024.

The reports of Blank Cycle, Repeatability and Calibration Values were all found in acceptable range.

Next calibration cycle is due on 18th June 2025.



Shrish Dixit
(Head- Products & Marketing)
For Horiba India Pvt. Ltd.

Explore the Future

Automotive Test Systems | Process & Environment | Medical | Semiconductor | Scientific

HORIBA
EQUIPMENT