



A A CALIBRATION PVT. LTD.

CALIBRATION CERTIFICATE

7.8F-01



CC - 2646

Certificate No	AACPL / 03400F	Field	Elec. Tech.	Group	Page No.	1 of 1
SRF No & Date	240506.1	DT - 06 / 05 / 2024		ULR No.	CC2646240000 03400F	
Calibration Method :	Direct Method			Accuracy		
Company Name	M/s Krsnaa Diagnostics Ltd.			Calibration Date	06.05.2024	
Address	Regional Hospital, Kotla Nala, Solan			Recommended Date for Next Calib.	05.05.2025	
				Certificate Issue Dt.	13.05.2024	
				UUC Condition	Good	

Calibration Instrument Detail

SI	Instrument Name	Range	Least Count	Make / SI.No.	Party ID / Location
1	Dig. Temp. Controller	2 to 8 °C	1 °C	LG MEZ64648238	KDHPRHSLBC10 Bio Chemistry (Refrigerator)

Standard Equipments Used (Traceable to National Standard)

SI	Instrument Name	Make/SI No	Calibrated By	Cal Certificate No	Cal. Valid Upto
1	Universal Calibrator	Radix / 112004873	AACPL	AACPL/02470F	01/04/2025

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
IS : 1248	Temp. Measurement	53 %RH	25.4 °C	Site

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 03, Section No : CPET - 01

Set Points :- 2, 5 & 8 °C

Calibration Results				
SI. No.	Std. Reading on Master in °C	UUC Average Reading in °C	Deviation in °C	Uncertainty at approx 95% confidence Level and coverage factor $k = 2$ is \pm
1	2.1	2	-0.1	1.62 °C
2	5.1	5	-0.1	1.62 °C
3	8.1	9	0.9	1.62 °C

Note

- 1 The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- 2 This certificate cannot be reproduced except in full without our prior permission in writing
- 3 This certificate refers only to the particular item(s) submitted for calibration
- 4 UUC - Unit under calibration

----- End of Certificate -----

Calibrated by
MariK



Approved by
[Signature]
Anand Chaurasia
(CEO)