



# A A CALIBRATION PVT. LTD.

## CALIBRATION CERTIFICATE

7.8F-01



CC - 2646

Certificate No	AACPL / 03439F	Field	Mech.	Group	Volume	Page No.	1 of 1
SRF No & Date	240507.1 DT - 07 / 05 / 2024		ULR No.	CC2646240000 03439F			
Calibration Method :	Comparison Method			Accuracy			
Company Name Address	M/s Krsnaa Diagnostics Ltd. Regional Hospital, Kotla Nala, Solani			Calibration Date	07.05.2024		
				Recommended Date for Next Calib.	06.05.2025		
				Certificate Issue Dt.	11.05.2024		
				UUC Condition	Good		

### Calibration Instrument Detail

SI	Instrument Name	Range	Least Count	Make / SI.No.	Party ID / Location
1	Micro Pipette	20 to 200 µl	1 µl	Dragon Lab YE222AX0047995	KDHPRHSLSE06 Serology

### Standard Equipments Used ( Traceable to National Standard )

SI	Instrument Name	Make/SI No	Calibrated By	Cal Certificate No	Cal. Valid Upto
1	Weighing Balance	Radwag / 668683	Bluebox Techno	BBT/744/JAN/24	04/03/2025

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
ISO - 8655 - 6 - 2022	Volume Measurement	53 %RH	23.7 °C	Laboratory

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 01C, Section No : CP - 03

### Calibration Result

S. No	Nominal Value on UUC (in µl)	Average Volume at 27 °C (in µl)	Deviation (in µl)	Uncertainty at approx 95% confidence Level and coverage factor $k = 2$ is $\pm$ (in µl)
1	20	20.00851	-0.00851	0.25
2	100	100.42597	-0.42597	0.25
3	200	200.76259	-0.76259	0.25

#### Note

- 1 The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- 2 This certificate cannot be reproduced except in full without our prior permission in writing
- 3 This certificate refers only to the particular item(s) submitted for calibration
- 4 UUC - Unit under calibration

----- End of Certificate -----

*Ashish*  
Calibrated by



Approved by  
*Anand Chaurasia*  
Anand Chaurasia  
(CEO)