



A A CALIBRATION PVT. LTD.

CALIBRATION CERTIFICATE

7.8F-01



CC - 2646

Certificate No	AACPL / 03402F	Field	Mech.	Group	Speed	Page No.	1 of 1
SRF No & Date	240506.1 DT - 06 / 05 / 2024			ULR No.	CC2646240000 03402F		
Calibration Method :	Comparison Method			Accuracy			
Company Name	M/s Krsnaa Diagnostics Ltd.			Calibration Date	06.05.2024		
Address	Regional Hospital, Kotla Nala, Solan			Recommended Date for Next Callb.	05.05.2025		
				Certificate Issue Dt.	13.05.2024		
				UUC Condition	Good		

Calibration Instrument Detail

SI	Instrument Name	Range	Least Count	Make / SI.No.	Party ID / Location
1	Centrifuge M/c	0 - 6000 RPM	10 RPM	Remi ZIHN-34769	KDHPRHSLCC12 Clinical Chemistry

Standard Equipments Used (Traceable to National Standard)

SI	Instrument Name	Make/SI No	Calibrated By	Cal Certificate No	Cal. Valid Upto
1	Dig. Tachometer	Line Seiki / -----	AACPL	AACPL / 06363F	31/08/2024

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
IS : 12508 - 2018	RPM Measurement	53 %RH	24.6 °C	Site

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 01B, Section No : CP - 08

Calibration Result

SI	Calib. Gauge Value	Master Gauge Value	Deviation	Uncertainty at approx 95% confidence Level and coverage factor $k = 2$ is \pm (in RPM)
	UUC (in RPM)	(in RPM)	(in RPM)	
1	100	99.8	0.2	5.88
2	500	499.5	0.5	5.88
3	1000	999.1	0.9	5.88
4	2500	2498.4	1.6	6.15
5	3500	3498.2	1.8	6.15
6	5000	4997.7	2.3	6.15

Note

- 1 The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- 2 This certificate cannot be reproduced except in full without our prior permission in writing
- 3 This certificate refers only to the particular item(s) submitted for calibration
- 4 UUC - Unit under calibration

----- End of Certificate -----


Calibrated by



Approved by

Anand Chaurasia
(CEO)