



APPENDIX E

THERMOMETER CALIBRATION SHEET

Calibration of Thermometer

Name and ID of equipment: Refrigerator-01 (PPTCT Hindu Rao Hospital)

ID of Thermometer to be calibrated: ICTC.T-1

Temperature range of thermometer: -10°C to +110°C

Name and ID of Calibrator: Cal T-2

Temperature range of Calibrator: -10°C to +110°C

Thermometer calibration done on: 11/07/2024

Thermometer calibration due on: 10/07/2025

Thermometer calibration done at: HIV Lab, MAMC

Time of reading	Temperature of Working Thermometer	Temperature of Calibrator	Difference in readings
After 1 hour	4°C	4°C	0
After 2 hour	4°C	4°C	0
After 3 hour	4°C	4°C	0

Correction Factor: 0

Comments: Nil

Calibrated By: Ramesh kumar

*Egman*

*Ramesh*  
11/07/2024

Technical Officer  
HIV LABORATORY  
Department of Microbiology  
Maulana Azad Medical College

Document Type: SOP for operation and monitoring of Refrigerator	Document No: MAMC/HIV/SOP/11-11000	
Issue No.:02	Issue Date: 10.09.2016	Page No.: 1 of 1