



ASIAN TECHNOLOGY

(A HOUSE OF CALIBRATION)



CALIBRATION CERTIFICATE

Customer name And Address	M/S: QPLUS Diagnostics Shop No.4, Pkt-2, Near Aandhra Bank, Opp, Royal Garden, Sec-24, Rohini, Delhi-85	Service request No. & date	P/02/14/08/24&14/08/2024
		ULR no.	CC401224000000272 F
		Cert. No.	AT/24000000272
		Date of Receipt of DUC	14/08/2024
		Date of calibration	14/08/2024
		Date of issue	16/08/2024
		Suggested due date	13/08/2025

Instrument Details			
Instrument name	Mini Centrifuge	I.D. / S. No.	MC-01
Make / Model No.	----	Location	Lab
Range / Size	0 to 3500 rpm	Calibration Performed at	Site
Least Count	10 rpm	Visual Inspection	OK

Detail of reference standards & Major equipments used			
Equipment Name	Digital Tachometer		
Make	LT Lutron		
Model / SR No.	-----/A.C92792		
Certificate No.	QCTS/00053/10		
Calibration Validity	19/10/2024		
Calibration by	Quality Calibration Testing Sol.		

Environmental Condition	Temperature	20±2 °C	Calibration Reference	IS: 12508
	Relative Humidity	50±10 %	Calibration Procedure	CP-37

Calibration Results			
Serial No.	Standard value in (rpm)	Measured value in (rpm)	Uncertainty At 95% C.L. (coverage factor K=2)
01.	503.5	500	±0.2 %rdg.
02.	1002	1000	
03.	2006	2000	
04.	3353	3348	

Remarks:

❖	(1) Standard equipment use for calibration are traceable to national/ international standards.
❖	(2) The reported expended uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by coverage factor = 2 such that the coverage probability corresponds to approximately 95%. (3) The above results are valid at the time of and under the stated conditions measurement. (4) This certificate is refers only to the particular item submitted for calibration. (5) Next calibration due date given as requested by the customer.

Calibrated By
(Calibration Engg./TM)
(M. H. PAL)
Form No.-QF-47

Approved By
(QM/TM)
(NEERAJ TYAGI)
Page No. 1 of 1



---End of Report---