

AGD Biomedicals (P) Ltd.
Address- Mehta Trade Center, Sir M.V.Road ,
Andheri East, Mumbai-400099
Tel – +91-22-28231061/66
Email : sales@agdbio.com

DOC.NO.- CC/AGD 2020/CCNR/03/00
DATE: 29 June 2024

CALIBRATION CERTIFICATE

This is to certify that the AGD 2020 Clinical Chemistry Analyzer bearing serial number: AO2107009 installed at CHETHANA CENTRE FOR NEUROPHYSIATRIC REHABILITATION MALAPARAMBA , CALICUT, KERALA is company Calibrated .

Verified with Quality control material Randox serum Control (Normal ,Lot no.1665UN)
The reports of Control values were all found in acceptable range.
Data given to laboratory.

Instrument is ready for routine work.

Please refer our Service Report on dated 28 June, 2024 for your reference.

Calibration Date: 28 June, 2024
Next Calibration Due on: 27 June, 2025

PHOTOMETER CALIBRATION
Cuvette Type- Flowcell
Status- PASS

Wavelength	Gain	Count	Factor	
1	340	111	23534	8
2	405	8	19910	7
3	450	8	31285	6
4	505	1	21073	8
5	546	1	24138	6
6	578	1	23868	6
7	630	1	35880	5
8	670	1	33074	5

Enclosed :

- 1) QC Run Data
- 2) Assay Sheet of Control
- 3) Printouts

Authorized Signatory

For AGD BIOMEDICALS Pvt. Ltd.