

Date: **23.07.2024**

CALIBRATION CERTIFICATE

This is to state that the hematology cell counter model HORIBA Medical: **YUMIZEN H550** bearing serial no: **302YADH05317** installed AT **LEGIT MEDICAL DIAGNOSTICS & POLYCLINIC, NIZAMABAD** calibrated with Calibrator on **23rd July, 2024**, Next calibration Due Date: - **22/07/2025**.

Subsequently controls were processed and found them in respective range.

Calibrator Used: ABX MINOCAL
Lot No.: CX 496
Expiry Date: 05/08/2024

Thanking you,

Yours truly,
For HORIBA MEDICAL



RAJU DASARI
ENGINEER CUSTOMER SUPPORT.