



PROFICIENCY TESTING REPORT
ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME
 NABL accredited program as per ISO/IEC 17043:2010 standard
 Organized By Department of Hematology, AIIMS, New Delhi-110029



Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens

EQAP CODE No. : 6026

Distribution No.: 163-0

Month/Year: April/2024

Instrument ID: Transasia

Model Name.: H560

Serial No.: K110432323123

Name & Contact No. of PT Co-ordinator: Dr. Manoranjan Mahapatra (Prof. & Head), Hematology, AIIMS, Delhi,
 Tel: 9013085730 , E-Mail : info@ishtmaiimseqap.com

Date of issue & status of the report: 05-06-2024[Final].

CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 ³ /µl	1	4.8	4.78	9.58	9.92	0.029	-0.46	0.02	0.1	0.007	-0.72
RBC x10 ⁶ /µl	1	5.58	5.57	11.15	10.24	0.014	2.46	0.01	0.05	0.003	-0.77
Hb g/dl	1	15.7	15.6	31.3	29.8	0.035	1.69	0.1	0.1	0.009	0.00
HCT%	1	50.3	50.2	100.5	90.6	0.204	1.79	0.1	0.4	0.028	-0.55
MCV-fl	1	90.2	90.2	180.4	177.3	0.277	0.41	0	0.3	0.020	-1.01
MCH-Pg	1	28.2	28.1	56.3	58.3	0.080	-0.96	0.1	0.2	0.016	-0.34
MCHC-g/dl	1	31.2	31.1	62.3	65.85	0.149	-0.92	0.1	0.3	0.020	-0.67
Plt. x10 ³ /µl	1	213	198	411	411	2.173	0.00	15	6	0.417	1.31
Retic %	2	2.8	2.7	5.5	14.45	0.187	-1.63	0.1	0.5	0.035	-0.67

P.S . Assesment

YOUR REPORT			CONSENSUS REPORT		
DLC%	3	Nrbcs=0 , Poly=27 L=1, E=1, Mono/Promono=0 , B1=3 P.M.=3, Mye=28, Meta=37, Other=0	Blast: 20-85, Poly: 2-8, Lympho: 2-7, Promyelo: 0-22, Myelo: 0-8, Meta: 0-7, nRBC/Mono/Eos/Baso/: 0-5		
RBC Morphology	3	Mild microcytic and hypochromic with mild anisopoikilocytosis	Predominantly: Normocytic/Normochromic, Mild: Anisocytosis		
Diagnosis	3	Smear suggestive of myeloproliferative disorder probably Chronic myeloid leukemia	Acute Promyelocytic Leukemia(APML)		

COMBINED DATA VALUES OF TOTAL PARTICIPANTS

Test parameters	S.No.	Total participants covered in the current dist. 163--O	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
WBC x10³/µl	1	324	324	84.57	83.95	3.4	6.48	12.03	9.57
RBC x10⁶/µl	1	324	324	87.35	88.27	5.56	5.56	7.09	6.17
Hb g/dl	1	324	324	89.2	91.67	6.79	3.09	4.01	5.24
HCT%	1	324	324	91.98	88.27	3.7	5.25	4.32	6.48
MCV-fl	1	324	324	91.98	84.88	4.01	7.41	4.01	7.71
MCH-Pg	1	324	324	87.35	91.67	7.72	2.16	4.93	6.17
MCHC-g/dl	1	324	324	90.43	87.35	5.56	4.63	4.01	8.02
Plt. x10³/µl	1	324	324	88.27	89.51	5.86	5.25	5.87	5.24
ReticCount%	2	324	256	93.75	83.2	4.3	13.28	1.95	3.52
PS Assessment	3	324	254	Satisfactory :93.83%, Borderline Sat. :2.16%, Unsatisfactory :4.01%					

***Comments:**

1). Among Lab (EQA) : PS Diagnosis partially correct, remaining results acceptable

2). Within Lab (IQA) : Precision acceptable.

Note-1: EQA (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

IQA (Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

Note-2: Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

Note-3: Z score 0 to ± 2 : Acceptable, Z score ± 2 to ± 3 :Warning Signal, Z score $> \pm 3$: Unacceptable [As per ISO/IEC 13528:2015 standard]

Note-4: Z score value between " 0 to ± 2 " are texted in green colour. Z score value between " ± 2 to ± 3 " are texted in orange colour. Z score value $> \pm 3$ are texted in red colour.

Note-5: Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value ($0.3*SDPA$). To pass the stability test, average difference in measurement values of first and last day sample ($\bar{x}-\bar{y}$) should be smaller than the check value ($0.3*SDPA$).

Note-6: ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

Note-7: Participants are free to use methods/analyzer of their own choice.

Note-8: Proficiency testing (PT) samples are sent quarterly to each participant.

Note-9: All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website www.ishtmaiimseqap.com.

Note 10: Reports are kept confidential.

Report authorized by,



Dr. Manoranjan Mahapatra (Prof. & Head)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----