

TESTING MACHINE SERVICE CENTRE

Calibration, Repairing, Servicing of All Type of Testing Machines, Chemical Lab Instruments Civil Lab Instruments, Temperature Measuring Instruments, Dimentional Instruments, Survey Instruments etc.





BALITIKURI SHIBTALA INDUSTRIAL ESTATE, HOWRAH - 711113, WB, INDIA

Website - www.testingmachine.org

CALIBRATION CERTIFICATE

PAGE: ONE OF ONE

SSUED TO:

M/S Govt. of West Bengal

Department of Health & Family Welfare Office of the

superintendent Walsh Sub-Divisional Hospital

Add: 22A, TC, Goswami Street, Serampore,

Hooghly - 712 201.

FICATE NO.

TMSC / JUN - 24 / 14 / 05

CERTIFICATE ISSUE DATE :

17-06-2024

ERATION DATE

14-06-2024

SUGGESTED CALIBRATION DUE DATE

13-06-2025

TON OF THE INSTRUMENT:

MENT NAME

Centrifuge (Analogue)

RANGE

MODEL / MODEL

Remi / ----

LEAST COUNT

FICATION NO.

WSDH/CEN-02

SERIAL NO.

: 2000059215

REMENT TRACEABILITY & ITS REFERENCE CERTIFICATE NO. & DATE:

lisstrument	Make	ID No.	Range	Certificate No	Validity	Calibrated by
Tachometer	METRIX	2.5-99999 rpm	TMSC/DT-01	TMSC/23/M-24937	28.12.2024	NABL LAB CC-2498

manmental Conditions: Temp. 25.0 To 30.0°C mation Method: Comparison Method

Relative Humidity:

40%-60%

BRATION RESULTS:

Sr. No.	Nominal value on UUC (rpm)	Average value of Std. (rpm)	Error (rpm)	Uncertainty of measurement at 95% confidence level with k=2 (±) (rpm)	
1	200	199.23	0.77	8.00	
2	500	498.19	1.81	8.00	
3	1000	995.01	4.99	14.00	
4	4000	3992.84	7.16	18.00	
5	5000	4980.34	19.66	18.00	

-The UUC* (Unit Under Calibration) has been calibrated for its range.

- In the certificate refers only to the particular item (s) submitted for calibration.
- This certificate refers to the values obtained at the time of calibration.
- This certificate shall not be reproduced, except in full, unless written permission has been obtained from CALIBRATION ENGINEER /TMSC.





(CALIBRATION ENGINEER) PIJUS ADHIKARY

Contact No: 9038033905, 9831233902 Email: testingmachine3@gmail.com