

TESTING MACHINE SERVICE CENTRE

ISO 9001-2015

Calibration, Repairing, Servicing of All Type of Testing Machines, Chemical Lab Instruments Civil Lab Instruments, Temperature Measuring Instruments, Dimentional Instruments, Survey Instruments etc. BALITIKURI SHIBTALA INDUSTRIAL ESTATE, HOWRAH - 711113, WB, INDIA





Website - www.testingmachine.org

CALIBRATION CERTIFICATE

PAGE: ONE OF ONE

SSUED TO:

M/S Govt. of West Bengal

Department of Health & Family Welfare Office of the

superintendent Walsh Sub-Divisional Hospital

Add: 22A, TC, Goswami Street, Serampore,

Hooghly - 712 201.

FICATE NO.

TMSC/JUN-24/14/04

CERTIFICATE ISSUE DATE :

17-06-2024

ERATION DATE

14-06-2024

SUGGESTED CALIBRATION DUE DATE :

13-06-2025

TION OF THE INSTRUMENT:

MENT NAME

Centrifuge (Digital)

RANGE

: 0 to 5250 rpm

MODEL MODEL

Remi / R-8C

LEAST COUNT

FICATION NO.

WSDH/CEN-01

1 rpm

EXECUTE: TRACEABILITY & ITS REFERENCE CERTIFICATE NO. & DATE:

Instrument	Make	ID No.	Range	Certificate No	Validity	Calibrated by
mail Tachometer	METRIX	2.5-99999 rpm	TMSC/DT-01	TMSC/23/M-24937	28.12.2024	NABL LAB CC-2498

mmental Conditions: Temp. 25.0 To 30.0°C

Relative Humidity:

40%-60%

mation Method: Comparison Method

EXAMPLE 2 :

Sr. No.	Nominal value on UUC (rpm)	Average value of Std. (rpm)	Error (rpm)	Uncertainty of measurement at 95% confidence level with k=2 (±) (rpm)
1	200	199.51	0.49	8.00
2	500	499.00	1.00	8.00
3	1000	995.89	4.11	14.00
4	4000	3992.11	7.89	18.00
5	5200	5180.02	19.98	18.00

- The UUC* (Unit Under Calibration) has been calibrated for its range.

- 1) This certificate refers only to the particular item (s) submitted for calibration.
- This certificate refers to the values obtained at the time of calibration.
- In this certificate shall not be reproduced, except in full, unless written permission has been obtained from CALIBRATION ENGINEER /TMSC.





APPRO♥ (CALIBRATION ENGINEER) PIJUS ADHIKARY

Contact No: 9038033905, 9831233902 Email: testingmachine3@gmail.com