

Sensa Core Medical Instrumentation Pvt Ltd

Plot No.3,EPIP Zone,Pashamylaram, Medak Distric-502307. www.sensacore.com

Instrument Name : Electrolyte Analyzer - ST-200 CL PLUS

Date of Testing : 08/11/2023

Instrument Number : ST - 10798

(Valid for One Year)

Calibration Slope :

S.NO	Na+	K+	iCa++	Cl
Mean	55.66	63.01	--	55.78
Ranges	35-85 mV/decade	35-85 mV/decade	--	35-85 mV/decade
Status	OK	OK	--	Ok

Quality Control (I) :

S.NO	Na+	K+	iCa++	Cl
Mean	120.2	1.86	--	87.3
Ranges	112.2 – 130.0 mmol/L	1.55 – 2.55 mmol/L	--	73.0 – 91.0 mmol/L
Status	OK	OK	--	OK

Quality Control (II) :

S.NO.	Na+	K+	iCa++	Cl
Mean	144.2	3.87	--	107.5
Ranges	133.0 – 147.0 mmol/L	3.51 – 4.51 mmol/L	--	94.0 – 108.0 mmol/L
Status	OK	OK	--	OK

Quality Control (III) :

S.NO.	Na+	K+	iCa++	Cl
Mean	163.8	6.20	--	125.5
Ranges	152.0 – 170.0 mmol/L	5.52 – 6.52 mmol/L	--	113.0 – 131.0 mmol/L
Status	OK	OK	--	OK

For

YUCCA DIAGNOSTICS



Authorized signatory with Stamp

Date : 08/11/2023