

ADVANCE CALIBRATION SERVICES



(ISO: 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob.: +91-9891121066, 9868309579 • E-mail: advance.cal@gmail.com

Calibration Certificate

CERTIFICATE NO. : 230821A/N-			Page :1	of 1
CALIBRATED ON. : 21 / 08 / 202			Ref. Std. : IS	5:12508:1988
CALIBRATION DUE. : 20 / 08 / 202			C.P. No. : A	CS/CP/10
SRF No.& Dt.OF RECPT : 230821A/N D		Phy. Inspection : O.K.		
ULR No. & ISSUE DATE : CC34272300		128F Dt. 23.08.2023	Cal. At	Site / Lab
CALIBRATED FO		Microbiology. ollege, AMU Aligarh -2	02002	
ITEM CALIBRAT	ED : Centrifuge			
Make		emi	Range : (0 to 3500) RPM	
L. Count		0 RPM	ID No. : Cent-01	
Location		st Floor	Deptt. : Microbiology	
ENVIRONMENT	CONDITION: Temp. { 30°C	C ± 10°C} R.H. {30%	% to 80%}	
STANDARD INS	TRUMENTS USED :			
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Tachometer	ACS/DTM/01/ 211-0155V	Godrej	C-220912-6-1	12/09/2023
RESULTS	: Mechanical D	Discipline		
* U. C. INSTT.		tandard Measured Error		
READS (RPM)		(RPM)	(RPM)	
1000		1006	-6 N SE	
2000		2008	-8	
3000		3012	-12	
3500		3514	-14	
	(average of readings)			

REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is \pm 3.60 rpm of reading at coverage Factor k=1.96.

*U.C. : Under Calibration.

- 1. Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- 2. The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- 3. This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- 4. The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- 5. Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- 6. Confirmity statement not provided since not required.



