



'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob.: +91-9891121066, 9868309579 • E-mail: advance.cal@gmail.com

Calibration Certificate

CERTIFICATE NO. : 230821A/N-10			Page :1	of 1	
CALIBRATED ON. : 21 / 08 / 2023			Ref. Std. : IS	5:12508:1988	
CALIBRATION DUE. : 20 / 08 / 2024			C.P. No. : A	CS/CP/10	
SRF No.& Dt.OF RECPT : 230821A/N D		8.2023	Phy. Inspection : O.K.		
ULR No. & ISSUE DATE : CC342723000		129F Dt. 23.08.2023	Cal. At	Site / Lab	
CALIBRATED F		Microbiology. ollege, AMU Aligarh -2	02002		
ITEM CALIBRAT					
Make		Remi		: (0 to 3500) RPM	
L. Count		RPM	ID No. : Cent-02		
Location		Floor	Deptt. : Microbiology		
ENVIRONMENT	CONDITION: Temp. { 30°C	± 10°C} R.H. {30%	% to 80%}		
STANDARD INS	TRUMENTS USED :	<u> </u>			
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE	
Tachometer	ACS/DTM/01/ 211-0155V	Godrej	C-220912-6-1	12/09/2023	
RESULTS	: Mechanical D	iscipline			
		indard Measured	Error		
READS (RPM)		(RPM)	(RPM)		
	1000	1002	-2		
2000		2006	E 6 10 EER		
3000		3010	-10		
3500		3510	-10		
	(average of readings)	SSP JAN	CP OR		

REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is ± 3.60 rpm of reading at coverage Factor k=1.96

Under Calibration.

Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.

The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
 This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.

4. The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.

5. Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade /

6. Confirmity statement not provided since not required.

