





'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob.: +91-9891121066, 9868309579 • E-mail: advance.cal@gmail.com

Calibration Certificate

CERTIFICATE NO. : 230821A/N-102			Page	: 1 of 1	
CALIBRATED ON. : 21 / 08 / 2023			Ref. Std.	: CG-11	
CALIBRATION DUE. : 20 / 08 / 2024			C.P. No.	: ACS/CP/07	
SRF No.& Dt.OF RECPT : 230821A/N Dt. 21.08.2023			Phy. Inspection : O.K.		
ULR No. & ISSUE DATE : CC3427230000		01123F Dt. 23.08.202	23 Cal. At	: Site / Lab	
CALIBRATED FOR	: M/s SRL, Deptt. o J.N. Medical	of Microbiology. College, AMU Aligarh	1 -202002		
ITEM CALIBRATED	: Digital Temperatu	re Indicator of Refrig	erator		
	Make :		Range : (-20 to 50) °C		
710	L. Count : 0.1	°C	ID No. : DTI-02		
	Location : 1 st	Floor	Deptt. : Microl	Deptt. : Microbiology	
ENVIRONMENT COM	NDITION: Temp. { 30	°C <u>+</u> 10°C} R.H. {	30% to 80%}		
STANDARD INSTRU	MENTS USED:	***************************************		55 .CF	
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE	
Universal Calibrator	ACS/UC/01 / 300561	Anshaankan	AIPL/2022/0846	10/11/2023	
RESULTS :	ET Discipline				
* U. C. INSTT.		STANDARD	ERROR		
READS (°C)		READING (°C)	(°C)		
2.0		2	CALIDRATI		
4.0		4	CE ALIB ATIC SEC		
6.0		6	ECT IBRITION ER		
8.1		8	0.1		
	(ave	rage of readings)			

REMARKS

The Expanded uncertainty of measurement at 95% confidence level is ± 1.35°C at coverage Factor k=1.96

*U.C.

Under Calibration.

Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.

The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above

This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.

The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
 Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use

6. Confirmity statement not provided since not required.



