





'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob.: +91-9891121066, 9868309579 • E-mail: advance.cal@gmail.com

Calibration Certificate

CERTIFICATE NO.	: 230821A/N-103		Page	: 1 of 1	
CALIBRATED ON. : 21 / 08 / 2023				; CG-11	
CALIBRATION DUE. : 20 / 08 / 2024			C.P. No.	: ACS/CP/07	
SRF No.& Dt.OF RECPT : 230821A/N Dt. 21.08.2023			Phy. Inspection : O.K.		
ULR No. & ISSUE DATE : CC34272300000		01124F Dt. 23.08.202	23 Cal. At	: Site / Lab	
CALIBRATED FOR	: M/s SRL, Deptt. o J.N. Medical (f Microbiology. College, AMU Aligarh	1-202002		
ITEM CALIBRATED	: Digital Temperatu	re Indicator of Refrig	erator		
Make			Range : (-20 to 50) °C		
L. Count		°C	ID No. : DTI-0		
Location		Floor	Deptt. : Microbiology		
ENVIRONMENT COM	NDITION: Temp. { 30	°C ± 10°C} R.H. {	30% to 80%}		
STANDARD INSTRU	MENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE	
Universal Calibrator	ACS/UC/01 / 300561	Anshaankan	AIPL/2022/0846	10/11/2023	
RESULTS :	ET Discipline				
* U. C. INSTT.		STANDARD	ERROR		
READS (°C)		READING (°C)	(°C)		
THE CALL	.0	2 1014	CALTARATI		
4.1		4	0.1		
6.1		6	0.1		
6 8.12 A		8	0.1		
	(ave	rage of readings)			

REMARKS

: The Expanded uncertainty of measurement at 95% confidence level + 1.35°C at coverage Factor k=1.96

Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.

 The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above. The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample idea
 This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
 The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.

- 5. Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for
- 6. Confirmity statement not provided since not required

